

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0062023.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 21 10:01:56 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0095830.D 1000 =P0095828.D 750 =P0095829.D
 250 =P0095831.D 50 =P0095832.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.090	3.007	3.039	3.102	2.816	3.011	E6	3.84
2) SA	Decachlorobiphenyl	2.209	2.060	2.117	2.276	2.036	2.140	E6	4.73
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	5.234	4.872	5.041	5.443	4.721	5.062	E4	5.65
22) L5	AR-1248-2	8.447	7.729	8.021	8.905	8.065	8.234	E4	5.51
23) L5	AR-1248-3	8.903	8.215	8.512	9.139	8.267	8.607	E4	4.68
24) L5	AR-1248-4	8.581	7.897	8.245	8.713	8.120	8.311	E4	4.02
25) L5	AR-1248-5	8.514	7.773	8.153	8.833	8.757	8.406	E4	5.26
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0095830.D 1000 =P0095828.D 750 =P0095829.D
 250 =P0095831.D 50 =P0095832.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.408	1.377	1.389	1.396	1.272	1.368	E6	4.03
2) SA	Decachlorobiphenyl	1.227	1.140	1.172	1.285	1.303	1.226	E6	5.74
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0062023.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 21 10:01:56 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0095830.D 1000 =P0095828.D 750 =P0095829.D
 250 =P0095831.D 50 =P0095832.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	2.780	2.626	2.722	2.882	2.550	2.712 E4	4.78
22) L5	AR-1248-2	3.956	3.814	3.808	4.410	4.006	3.999 E4	6.15
23) L5	AR-1248-3	4.374	3.994	4.156	4.614	4.375	4.303 E4	5.50
24) L5	AR-1248-4	5.366	4.653	5.101	5.626	5.261	5.201 E4	6.95
25) L5	AR-1248-5	5.010	4.324	4.785	5.295	5.670	5.017 E4	10.16
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range