

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062023\
 Data File : P0095847.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2023 13:56
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO062023AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 14:45:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.417	3.637	151.5E6	69468186	52.078	51.186
2)	SA Decachlor...	10.236	8.673	108.8E6	60596151	52.509	49.418
Target Compounds							
3)	L1 AR-1016-1	5.594	4.725	958377	500968	11.613	10.962
4)	L1 AR-1016-2	5.625	4.743	1421421	383467	11.929	6.152 #
5)	L1 AR-1016-3	5.682	4.922	781705	241378	10.132	7.165 #
6)	L1 AR-1016-4	5.780	4.965	672296	15068994	11.217	457.364 #
7)	L1 AR-1016-5	6.077	5.178	14999553	8771348	241.390	240.683
8)	L2 AR-1221-1	4.613	3.802f	115377	198500	3.226	11.547 #
9)	L2 AR-1221-2	4.725	3.937	2037889	1046889	76.619	79.380
10)	L2 AR-1221-3	4.791	0.000	230334	0	2.886	N.D. #
11)	L3 AR-1232-1	4.791	0.000	230334	0	3.917	N.D. #
12)	L3 AR-1232-2	5.326	4.743	458109	383467	13.170	12.901
13)	L3 AR-1232-3	5.625	4.922	1421421	241378	25.000	15.112 #
14)	L3 AR-1232-4	5.780	5.005	672296	4530535	23.482	288.460 #
15)	L3 AR-1232-5	5.871	5.178	28205670	8771348	1135.600	487.431 #
16)	L4 AR-1242-1	5.594	4.725	958377	500968	14.499	13.857
17)	L4 AR-1242-2	5.625	4.743	1421421	383467	14.908	7.817 #
18)	L4 AR-1242-3	5.682	4.922	781705	241378	12.321	9.002 #
19)	L4 AR-1242-4	5.780	5.005	672296	4530535	13.738	153.240 #
20)	L4 AR-1242-5	6.526	5.532	14228578	35024217	287.761	986.587 #
21)	L5 AR-1248-1	5.594	4.725	958377	500968	18.933	18.471
22)	L5 AR-1248-2	5.871	4.965	28205670	15068994	342.566	376.850
23)	L5 AR-1248-3	6.077	5.005	14999553	4530535	174.265	105.297 #
24)	L5 AR-1248-4	6.482	5.178	18522283	8771348	222.859	168.638
25)	L5 AR-1248-5	6.526	5.573	14228578	4117449	169.269	82.074 #
26)	L6 AR-1254-1	6.457	5.532	48252989	35024217	506.775	472.893
27)	L6 AR-1254-2	6.677	5.681	69756618	31819873	502.279	466.280
28)	L6 AR-1254-3	7.045	6.087	66935449	48533746	495.446	487.385
29)	L6 AR-1254-4	7.332	6.315	44093544	29083148	491.746	492.768
30)	L6 AR-1254-5	7.754	6.736	52455650	44201395	509.935	495.711
31)	L7 AR-1260-1	7.211	6.218	30922698	25753758	271.968	358.833 #
32)	L7 AR-1260-2	7.469	6.406	28663972	20457641	236.066	245.585
33)	L7 AR-1260-3	7.754	6.559	52455650	20762255	378.371	269.524 #
34)	L7 AR-1260-4	8.044	7.034	20215097	2439415	211.429	42.766 #
35)	L7 AR-1260-5	8.381	7.277	7507240	5435198	44.967	42.341
36)	L8 AR-1262-1	7.815	6.826	8716103	2079261	64.935	54.361
37)	L8 AR-1262-2	8.381	7.034	7507240	2439415	36.234	30.941
38)	L8 AR-1262-3	8.714	7.561	5859152	397943	38.710	6.421 #
39)	L8 AR-1262-4	8.803	7.624	55044	5407082	0.456	49.750 #
40)	L8 AR-1262-5	9.464	8.124	263337	229231	3.486	4.513 #
41)	L9 AR-1268-1	8.714	7.561	5859152	397943	21.428	2.251 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062023\
Data File : P0095847.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2023 13:56
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 14:45:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.803	7.624	55044	5407082	0.223	34.527 #
43) L9	AR-1268-3	9.032	7.836	152135	298223	0.669	2.140 #
44) L9	AR-1268-4	9.464	8.124	263337	229231	3.182	4.127 #
45) L9	AR-1268-5	9.882	8.415	594337	340152	0.797	0.813



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062023\
Data File : P0095847.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2023 13:56
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 14:45:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.417 min
Delta R.T.: 0.002 min
Response: 151521508
Conc: 52.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: 0.001 min
Response: 69468186
Conc: 51.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.009 min
Response: 108784373
Conc: 52.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.673 min
Delta R.T.: 0.003 min
Response: 60596151
Conc: 49.42 ng/ml



#3 AR-1016-1


R.T.: 5.594 min
Delta R.T.: 0.002 min
Response: 958377
Conc: 11.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#3 AR-1016-1

R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 500968
Conc: 10.96 ng/ml




#4 AR-1016-2

R.T.: 5.625 min
Delta R.T.: 0.010 min
Response: 1421421
Conc: 11.93 ng/ml



#4 AR-1016-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 383467
Conc: 6.15 ng/ml



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: 0.004 min
Response: 781705
Conc: 10.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#5 AR-1016-3

R.T.: 4.922 min
Delta R.T.: 0.002 min
Response: 241378
Conc: 7.16 ng/ml




#6 AR-1016-4

R.T.: 5.780 min
Delta R.T.: 0.004 min
Response: 672296
Conc: 11.22 ng/ml



#6 AR-1016-4


R.T.: 4.965 min
Delta R.T.: 0.002 min
Response: 15068994
Conc: 457.36 ng/ml



#7 AR-1016-5


R.T.: 6.077 min
Delta R.T.: 0.003 min
Response: 14999553
Conc: 241.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#7 AR-1016-5

R.T.: 5.178 min
Delta R.T.: 0.002 min
Response: 8771348
Conc: 240.68 ng/ml




#8 AR-1221-1

R.T.: 4.613 min
Delta R.T.: -0.010 min
Response: 115377
Conc: 3.23 ng/ml



#8 AR-1221-1


R.T.: 3.802 min
Delta R.T.: -0.048 min
Response: 198500
Conc: 11.55 ng/ml



#9 AR-1221-2


R.T.: 4.725 min
Delta R.T.: 0.014 min
Response: 2037889
Conc: 76.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#9 AR-1221-2

R.T.: 3.937 min
Delta R.T.: 0.002 min
Response: 1046889
Conc: 79.38 ng/ml




#10 AR-1221-3

R.T.: 4.791 min
Delta R.T.: 0.004 min
Response: 230334
Conc: 2.89 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.011 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.791 min
Delta R.T.: 0.004 min
Response: 230334
Conc: 3.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.011 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.326 min
Delta R.T.: 0.005 min
Response: 458109
Conc: 13.17 ng/ml



#12 AR-1232-2


R.T.: 4.743 min
Delta R.T.: 0.001 min
Response: 383467
Conc: 12.90 ng/ml



#13 AR-1232-3


R.T.: 5.625 min
Delta R.T.: 0.011 min
Response: 1421421
Conc: 25.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#13 AR-1232-3

R.T.: 4.922 min
Delta R.T.: 0.003 min
Response: 241378
Conc: 15.11 ng/ml




#14 AR-1232-4

R.T.: 5.780 min
Delta R.T.: 0.005 min
Response: 672296
Conc: 23.48 ng/ml



#14 AR-1232-4


R.T.: 5.005 min
Delta R.T.: 0.002 min
Response: 4530535
Conc: 288.46 ng/ml



#15 AR-1232-5


R.T.: 5.871 min
Delta R.T.: 0.004 min
Response: 28205670
Conc: 1135.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#15 AR-1232-5

R.T.: 5.178 min
Delta R.T.: 0.002 min
Response: 8771348
Conc: 487.43 ng/ml




#16 AR-1242-1

R.T.: 5.594 min
Delta R.T.: 0.003 min
Response: 958377
Conc: 14.50 ng/ml



#16 AR-1242-1


R.T.: 4.725 min
Delta R.T.: 0.001 min
Response: 500968
Conc: 13.86 ng/ml



#17 AR-1242-2


R.T.: 5.625 min
Delta R.T.: 0.011 min
Response: 1421421
Conc: 14.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#17 AR-1242-2

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 383467
Conc: 7.82 ng/ml




#18 AR-1242-3

R.T.: 5.682 min
Delta R.T.: 0.005 min
Response: 781705
Conc: 12.32 ng/ml



#18 AR-1242-3


R.T.: 4.922 min
Delta R.T.: 0.003 min
Response: 241378
Conc: 9.00 ng/ml



#19 AR-1242-4


R.T.: 5.780 min
Delta R.T.: 0.004 min
Response: 672296
Conc: 13.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#19 AR-1242-4

R.T.: 5.005 min
Delta R.T.: 0.000 min
Response: 4530535
Conc: 153.24 ng/ml




#20 AR-1242-5

R.T.: 6.526 min
Delta R.T.: 0.008 min
Response: 14228578
Conc: 287.76 ng/ml



#20 AR-1242-5


R.T.: 5.532 min
Delta R.T.: 0.003 min
Response: 35024217
Conc: 986.59 ng/ml



#21 AR-1248-1


R.T.: 5.594 min
Delta R.T.: 0.003 min
Response: 958377
Conc: 18.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#21 AR-1248-1

R.T.: 4.725 min
Delta R.T.: 0.002 min
Response: 500968
Conc: 18.47 ng/ml




#22 AR-1248-2

R.T.: 5.871 min
Delta R.T.: 0.003 min
Response: 28205670
Conc: 342.57 ng/ml



#22 AR-1248-2


R.T.: 4.965 min
Delta R.T.: 0.002 min
Response: 15068994
Conc: 376.85 ng/ml



#23 AR-1248-3


R.T.: 6.077 min
Delta R.T.: 0.003 min
Response: 14999553
Conc: 174.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#23 AR-1248-3

R.T.: 5.005 min
Delta R.T.: 0.001 min
Response: 4530535
Conc: 105.30 ng/ml




#24 AR-1248-4

R.T.: 6.482 min
Delta R.T.: 0.003 min
Response: 18522283
Conc: 222.86 ng/ml



#24 AR-1248-4


R.T.: 5.178 min
Delta R.T.: 0.003 min
Response: 8771348
Conc: 168.64 ng/ml



#25 AR-1248-5


R.T.: 6.526 min
Delta R.T.: 0.008 min
Response: 14228578
Conc: 169.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#25 AR-1248-5

R.T.: 5.573 min
Delta R.T.: 0.002 min
Response: 4117449
Conc: 82.07 ng/ml




#26 AR-1254-1

R.T.: 6.457 min
Delta R.T.: 0.003 min
Response: 48252989
Conc: 506.77 ng/ml



#26 AR-1254-1


R.T.: 5.532 min
Delta R.T.: 0.002 min
Response: 35024217
Conc: 472.89 ng/ml



#27 AR-1254-2


R.T.: 6.677 min
Delta R.T.: 0.003 min
Response: 69756618
Conc: 502.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#27 AR-1254-2

R.T.: 5.681 min
Delta R.T.: 0.002 min
Response: 31819873
Conc: 466.28 ng/ml




#28 AR-1254-3

R.T.: 7.045 min
Delta R.T.: 0.003 min
Response: 66935449
Conc: 495.45 ng/ml



#28 AR-1254-3


R.T.: 6.087 min
Delta R.T.: 0.003 min
Response: 48533746
Conc: 487.39 ng/ml



#29 AR-1254-4


R.T.: 7.332 min
Delta R.T.: 0.003 min
Response: 44093544
Conc: 491.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#29 AR-1254-4

R.T.: 6.315 min
Delta R.T.: 0.002 min
Response: 29083148
Conc: 492.77 ng/ml




#30 AR-1254-5

R.T.: 7.754 min
Delta R.T.: 0.003 min
Response: 52455650
Conc: 509.94 ng/ml



#30 AR-1254-5


R.T.: 6.736 min
Delta R.T.: 0.002 min
Response: 44201395
Conc: 495.71 ng/ml



#31 AR-1260-1


R.T.: 7.211 min
Delta R.T.: 0.003 min
Response: 30922698
Conc: 271.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#31 AR-1260-1

R.T.: 6.218 min
Delta R.T.: 0.003 min
Response: 25753758
Conc: 358.83 ng/ml




#32 AR-1260-2

R.T.: 7.469 min
Delta R.T.: 0.003 min
Response: 28663972
Conc: 236.07 ng/ml



#32 AR-1260-2


R.T.: 6.406 min
Delta R.T.: 0.001 min
Response: 20457641
Conc: 245.58 ng/ml



#33 AR-1260-3


R.T.: 7.754 min
Delta R.T.: 0.005 min
Response: 52455650
Conc: 378.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#33 AR-1260-3

R.T.: 6.559 min
Delta R.T.: 0.001 min
Response: 20762255
Conc: 269.52 ng/ml




#34 AR-1260-4

R.T.: 8.044 min
Delta R.T.: -0.010 min
Response: 20215097
Conc: 211.43 ng/ml



#34 AR-1260-4


R.T.: 7.034 min
Delta R.T.: 0.002 min
Response: 2439415
Conc: 42.77 ng/ml



#35 AR-1260-5


R.T.: 8.381 min
Delta R.T.: 0.003 min
Response: 7507240
Conc: 44.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#35 AR-1260-5

R.T.: 7.277 min
Delta R.T.: 0.001 min
Response: 5435198
Conc: 42.34 ng/ml




#36 AR-1262-1

R.T.: 7.815 min
Delta R.T.: -0.016 min
Response: 8716103
Conc: 64.93 ng/ml



#36 AR-1262-1


R.T.: 6.826 min
Delta R.T.: 0.000 min
Response: 2079261
Conc: 54.36 ng/ml



#37 AR-1262-2


R.T.: 8.381 min
Delta R.T.: 0.000 min
Response: 7507240
Conc: 36.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#37 AR-1262-2

R.T.: 7.034 min
Delta R.T.: 0.000 min
Response: 2439415
Conc: 30.94 ng/ml




#38 AR-1262-3

R.T.: 8.714 min
Delta R.T.: 0.014 min
Response: 5859152
Conc: 38.71 ng/ml



#38 AR-1262-3


R.T.: 7.561 min
Delta R.T.: 0.000 min
Response: 397943
Conc: 6.42 ng/ml



#39 AR-1262-4


R.T.: 8.803 min
Delta R.T.: 0.013 min
Response: 55044
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#39 AR-1262-4

R.T.: 7.624 min
Delta R.T.: -0.002 min
Response: 5407082
Conc: 49.75 ng/ml




#40 AR-1262-5

R.T.: 9.464 min
Delta R.T.: 0.004 min
Response: 263337
Conc: 3.49 ng/ml



#40 AR-1262-5


R.T.: 8.124 min
Delta R.T.: 0.002 min
Response: 229231
Conc: 4.51 ng/ml



#41 AR-1268-1


R.T.: 8.714 min
Delta R.T.: 0.016 min
Response: 5859152
Conc: 21.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#41 AR-1268-1

R.T.: 7.561 min
Delta R.T.: -0.001 min
Response: 397943
Conc: 2.25 ng/ml




#42 AR-1268-2

R.T.: 8.803 min
Delta R.T.: 0.010 min
Response: 55044
Conc: 0.22 ng/ml



#42 AR-1268-2


R.T.: 7.624 min
Delta R.T.: -0.004 min
Response: 5407082
Conc: 34.53 ng/ml



#43 AR-1268-3


R.T.: 9.032 min
Delta R.T.: 0.005 min
Response: 152135
Conc: 0.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254




#43 AR-1268-3

R.T.: 7.836 min
Delta R.T.: 0.000 min
Response: 298223
Conc: 2.14 ng/ml




#44 AR-1268-4

R.T.: 9.464 min
Delta R.T.: 0.004 min
Response: 263337
Conc: 3.18 ng/ml



#44 AR-1268-4


R.T.: 8.124 min
Delta R.T.: 0.000 min
Response: 229231
Conc: 4.13 ng/ml



#45 AR-1268-5

R.T.: 9.882 min
Delta R.T.: -0.005 min
Response: 594337
Conc: 0.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062023AR1254



#45 AR-1268-5

R.T.: 8.415 min
Delta R.T.: -0.002 min
Response: 340152
Conc: 0.81 ng/ml