

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062024\
 Data File : P0104377.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 17:09
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:40:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 21 01:40:08 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.405	3.548	208.2E6	39328833	25.318	25.171
2)	SA Decachlor...	10.105	8.462	150.7E6	85796016	26.174	26.389
Target Compounds							
16)	L4 AR-1242-1	5.565	4.606	62501146	9899765	270.378	266.167
17)	L4 AR-1242-2	5.587	4.624	87202990	16321662	269.524	262.683
18)	L4 AR-1242-3	5.649	4.796	60297441	8962710	286.674	255.750
19)	L4 AR-1242-4	5.747	4.879	46847282	7087972	275.280	260.258
20)	L4 AR-1242-5	6.480	5.393	47316157	9588140	272.965	260.103



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062024\
Data File : PO104377.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 17:09
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:40:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062024.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 21 01:40:08 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 208151108
Conc: 25.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.548 min
Delta R.T.: 0.000 min
Response: 39328833
Conc: 25.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.105 min
Delta R.T.: 0.000 min
Response: 150737681
Conc: 26.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.462 min
Delta R.T.: 0.000 min
Response: 85796016
Conc: 26.39 ng/ml



#16 AR-1242-1


R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 62501146
Conc: 270.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.606 min
Delta R.T.: 0.000 min
Response: 9899765
Conc: 266.17 ng/ml




#17 AR-1242-2

R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 87202990
Conc: 269.52 ng/ml



#17 AR-1242-2


R.T.: 4.624 min
Delta R.T.: 0.000 min
Response: 16321662
Conc: 262.68 ng/ml



#18 AR-1242-3


R.T.: 5.649 min
Delta R.T.: 0.000 min
Response: 60297441
Conc: 286.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.796 min
Delta R.T.: 0.000 min
Response: 8962710
Conc: 255.75 ng/ml




#19 AR-1242-4

R.T.: 5.747 min
Delta R.T.: 0.000 min
Response: 46847282
Conc: 275.28 ng/ml



#19 AR-1242-4


R.T.: 4.879 min
Delta R.T.: 0.000 min
Response: 7087972
Conc: 260.26 ng/ml



#20 AR-1242-5

R.T.: 6.480 min
Delta R.T.: 0.000 min
Response: 47316157
Conc: 272.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.393 min
Delta R.T.: 0.000 min
Response: 9588140
Conc: 260.10 ng/ml