

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062024\
 Data File : P0104382.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 18:32
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:58:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 21 01:47:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.405	3.548	206.7E6	39836103	25.687	25.385
2)	SA Decachlor...	10.104	8.463	151.2E6	86163585	26.332	26.521
Target Compounds							
21)	L5 AR-1248-1	5.565	4.606	49590113	7520541	283.263	266.016
22)	L5 AR-1248-2	5.838	4.838	69056548	9125114	279.951	268.530
23)	L5 AR-1248-3	6.041	4.879	69929659	10468379	269.285	265.902
24)	L5 AR-1248-4	6.442	5.047	78041192	13418948	269.426	267.333
25)	L5 AR-1248-5	6.481	5.433	75874227	15705394	269.798	266.399



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062024\
Data File : PO104382.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 18:32
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:58:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062024.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 21 01:47:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 206707621
Conc: 25.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.548 min
Delta R.T.: 0.000 min
Response: 39836103
Conc: 25.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.104 min
Delta R.T.: 0.000 min
Response: 151175536
Conc: 26.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.463 min
Delta R.T.: 0.000 min
Response: 86163585
Conc: 26.52 ng/ml



#21 AR-1248-1


R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 49590113
Conc: 283.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.606 min
Delta R.T.: 0.000 min
Response: 7520541
Conc: 266.02 ng/ml




#22 AR-1248-2

R.T.: 5.838 min
Delta R.T.: 0.000 min
Response: 69056548
Conc: 279.95 ng/ml



#22 AR-1248-2


R.T.: 4.838 min
Delta R.T.: 0.000 min
Response: 9125114
Conc: 268.53 ng/ml



#23 AR-1248-3


R.T.: 6.041 min
Delta R.T.: 0.000 min
Response: 69929659
Conc: 269.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 4.879 min
Delta R.T.: 0.000 min
Response: 10468379
Conc: 265.90 ng/ml




#24 AR-1248-4

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 78041192
Conc: 269.43 ng/ml



#24 AR-1248-4


R.T.: 5.047 min
Delta R.T.: 0.000 min
Response: 13418948
Conc: 267.33 ng/ml



#25 AR-1248-5

R.T.: 6.481 min
Delta R.T.: 0.000 min
Response: 75874227
Conc: 269.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.433 min
Delta R.T.: 0.000 min
Response: 15705394
Conc: 266.40 ng/ml