

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0062025\
 Data File : P0111760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2025 15:27
 Operator : YP/AJ
 Sample : Q2126-07
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 00:05:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.677	3.673	116.8E6	102.4E6	20.286	18.223
2)	SA Decachlor...	8.712	8.659	101.6E6	35220630	19.363	19.813
Target Compounds							
3)	L1 AR-1016-1	4.765	4.749	6645519	5074942	27.858	25.792
4)	L1 AR-1016-2	4.784	4.768	6337050	5205595	19.045	18.273
5)	L1 AR-1016-3	4.840	4.944	3522410	2972788	15.305	19.787 #
6)	L1 AR-1016-4	4.961	4.986	4483657	6953295	24.176	57.094 #
7)	L1 AR-1016-5	5.217	5.197	10609398	7867836	56.843	49.422
8)	L2 AR-1221-1	3.880	3.888	57214	110682	0.657	1.482 #
9)	L2 AR-1221-2	3.980	3.970	2490450	1851881	38.812	33.100
10)	L2 AR-1221-3	4.061	4.043	543683	289409	2.788	1.684 #
11)	L3 AR-1232-1	4.061	4.043	543683	289409	3.583	2.176 #
12)	L3 AR-1232-2	4.542	4.768	3505814	5205595	40.661	39.499
13)	L3 AR-1232-3	4.784	4.944	6337050	2972788	38.906	43.010
14)	L3 AR-1232-4	4.961	5.027	4483657	5962342	51.525	98.577 #
15)	L3 AR-1232-5	5.003	5.197	7915685	7867836	139.901	118.465
16)	L4 AR-1242-1	4.765	4.749	6645519	5074942	32.989	31.181
17)	L4 AR-1242-2	4.784	4.768	6337050	5205595	22.490	22.052
18)	L4 AR-1242-3	4.840	4.944	3522410	2972788	17.716	23.671 #
19)	L4 AR-1242-4	4.961	5.027	4483657	5962342	28.342	48.649 #
20)	L4 AR-1242-5	5.613	5.548	11679560	10906692	69.847	72.765
21)	L5 AR-1248-1	4.765	4.749	6645519	5074942	41.209	38.076
22)	L5 AR-1248-2	5.003	4.986	7915685	6953295	36.714	39.029
23)	L5 AR-1248-3	5.217	5.027	10609398	5962342	39.332	31.474
24)	L5 AR-1248-4	5.571	5.197	17571892	7867836	43.951	34.459
25)	L5 AR-1248-5	5.613	5.590	11679560	7811426	41.592	35.871
26)	L6 AR-1254-1	5.571	5.548	17571892	10906692	41.339	33.766
27)	L6 AR-1254-2	5.721	5.696	9607870	4262534	25.315	15.261 #
28)	L6 AR-1254-3	6.125	6.097	15868688	6323016	27.774	15.089 #
29)	L6 AR-1254-4	6.349	6.321	11535842	2651152	33.065	11.638 #
30)	L6 AR-1254-5	6.768	6.741	5774076	943028	11.275	2.896 #
31)	L7 AR-1260-1	6.268	6.242	18265131	3232508	51.404	12.883 #
32)	L7 AR-1260-2	6.444	6.414	5225914	913976	11.061	3.109 #
33)	L7 AR-1260-3	6.811	6.565	10975960	624549	25.846	2.338 #
34)	L7 AR-1260-4	7.066	7.036	2690924	562199	8.558	2.957 #
35)	L7 AR-1260-5	7.295	7.282	8661041	844106	10.514	1.986 #
36)	L8 AR-1262-1	6.811	6.777	10975960	1243260	17.997	3.682 #
37)	L8 AR-1262-2	7.295	7.282	8661041	844106	9.200	1.812 #
39)	L8 AR-1262-4	7.657	7.624	409098	365923	0.623	1.321 #
40)	L8 AR-1262-5	8.177	8.118	215047	67570	0.706	0.665
42)	L9 AR-1268-2	7.657	7.624	409098	365923	0.419	0.879 #
43)	L9 AR-1268-3	7.850	7.817	476089	143675	0.581	0.453

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062025\
 Data File : P0111760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2025 15:27
 Operator : YP/AJ
 Sample : Q2126-07
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 00:05:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	8.177	8.118	215047	67570	0.625	0.584
45) L9	AR-1268-5	8.450	8.412	520140	308023	0.234	0.425 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062025\
Data File : P0111760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2025 15:27
Operator : YP/AJ
Sample : Q2126-07
Misc : AR1248 LOD 40 PPB
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 00:05:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.677 min
Delta R.T.: 0.000 min
Response: 116836317
Conc: 20.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#1 Tetrachloro-m-xylene

R.T.: 3.673 min
Delta R.T.: 0.000 min
Response: 102365496
Conc: 18.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.712 min
Delta R.T.: 0.000 min
Response: 101638831
Conc: 19.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.659 min
Delta R.T.: -0.002 min
Response: 35220630
Conc: 19.81 ng/ml



#3 AR-1016-1


R.T.: 4.765 min
Delta R.T.: -0.001 min
Response: 6645519
Conc: 27.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#3 AR-1016-1

R.T.: 4.749 min
Delta R.T.: -0.002 min
Response: 5074942
Conc: 25.79 ng/ml




#4 AR-1016-2

R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 6337050
Conc: 19.04 ng/ml



#4 AR-1016-2


R.T.: 4.768 min
Delta R.T.: -0.002 min
Response: 5205595
Conc: 18.27 ng/ml



#5 AR-1016-3


R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 3522410
Conc: 15.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#5 AR-1016-3

R.T.: 4.944 min
Delta R.T.: 0.000 min
Response: 2972788
Conc: 19.79 ng/ml




#6 AR-1016-4

R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 4483657
Conc: 24.18 ng/ml



#6 AR-1016-4


R.T.: 4.986 min
Delta R.T.: 0.000 min
Response: 6953295
Conc: 57.09 ng/ml



#7 AR-1016-5


R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 10609398
Conc: 56.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#7 AR-1016-5

R.T.: 5.197 min
Delta R.T.: -0.002 min
Response: 7867836
Conc: 49.42 ng/ml




#8 AR-1221-1

R.T.: 3.880 min
Delta R.T.: -0.009 min
Response: 57214
Conc: 0.66 ng/ml



#8 AR-1221-1


R.T.: 3.888 min
Delta R.T.: 0.005 min
Response: 110682
Conc: 1.48 ng/ml



#9 AR-1221-2


R.T.: 3.980 min
Delta R.T.: 0.004 min
Response: 2490450
Conc: 38.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#9 AR-1221-2

R.T.: 3.970 min
Delta R.T.: 0.003 min
Response: 1851881
Conc: 33.10 ng/ml




#10 AR-1221-3

R.T.: 4.061 min
Delta R.T.: 0.010 min
Response: 543683
Conc: 2.79 ng/ml



#10 AR-1221-3


R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 289409
Conc: 1.68 ng/ml



#11 AR-1232-1


R.T.: 4.061 min
Delta R.T.: 0.010 min
Response: 543683
Conc: 3.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#11 AR-1232-1

R.T.: 4.043 min
Delta R.T.: 0.000 min
Response: 289409
Conc: 2.18 ng/ml




#12 AR-1232-2

R.T.: 4.542 min
Delta R.T.: -0.001 min
Response: 3505814
Conc: 40.66 ng/ml



#12 AR-1232-2


R.T.: 4.768 min
Delta R.T.: -0.001 min
Response: 5205595
Conc: 39.50 ng/ml



#13 AR-1232-3


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 6337050
Conc: 38.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#13 AR-1232-3

R.T.: 4.944 min
Delta R.T.: 0.000 min
Response: 2972788
Conc: 43.01 ng/ml




#14 AR-1232-4

R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 4483657
Conc: 51.52 ng/ml



#14 AR-1232-4


R.T.: 5.027 min
Delta R.T.: -0.001 min
Response: 5962342
Conc: 98.58 ng/ml



#15 AR-1232-5


R.T.: 5.003 min
Delta R.T.: -0.001 min
Response: 7915685
Conc: 139.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#15 AR-1232-5

R.T.: 5.197 min
Delta R.T.: -0.002 min
Response: 7867836
Conc: 118.47 ng/ml




#16 AR-1242-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 6645519
Conc: 32.99 ng/ml



#16 AR-1242-1


R.T.: 4.749 min
Delta R.T.: -0.002 min
Response: 5074942
Conc: 31.18 ng/ml



#17 AR-1242-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 6337050
Conc: 22.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#17 AR-1242-2

R.T.: 4.768 min
Delta R.T.: -0.002 min
Response: 5205595
Conc: 22.05 ng/ml




#18 AR-1242-3

R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 3522410
Conc: 17.72 ng/ml



#18 AR-1242-3


R.T.: 4.944 min
Delta R.T.: 0.000 min
Response: 2972788
Conc: 23.67 ng/ml



#19 AR-1242-4


R.T.: 4.961 min
Delta R.T.: 0.001 min
Response: 4483657
Conc: 28.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#19 AR-1242-4

R.T.: 5.027 min
Delta R.T.: -0.001 min
Response: 5962342
Conc: 48.65 ng/ml




#20 AR-1242-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 11679560
Conc: 69.85 ng/ml



#20 AR-1242-5


R.T.: 5.548 min
Delta R.T.: -0.001 min
Response: 10906692
Conc: 72.77 ng/ml



#21 AR-1248-1


R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 6645519
Conc: 41.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#21 AR-1248-1

R.T.: 4.749 min
Delta R.T.: -0.001 min
Response: 5074942
Conc: 38.08 ng/ml




#22 AR-1248-2

R.T.: 5.003 min
Delta R.T.: 0.001 min
Response: 7915685
Conc: 36.71 ng/ml



#22 AR-1248-2


R.T.: 4.986 min
Delta R.T.: 0.000 min
Response: 6953295
Conc: 39.03 ng/ml



#23 AR-1248-3


R.T.: 5.217 min
Delta R.T.: 0.002 min
Response: 10609398
Conc: 39.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#23 AR-1248-3

R.T.: 5.027 min
Delta R.T.: 0.000 min
Response: 5962342
Conc: 31.47 ng/ml




#24 AR-1248-4

R.T.: 5.571 min
Delta R.T.: 0.002 min
Response: 17571892
Conc: 43.95 ng/ml



#24 AR-1248-4


R.T.: 5.197 min
Delta R.T.: -0.001 min
Response: 7867836
Conc: 34.46 ng/ml



#25 AR-1248-5


R.T.: 5.613 min
Delta R.T.: 0.002 min
Response: 11679560
Conc: 41.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#25 AR-1248-5

R.T.: 5.590 min
Delta R.T.: 0.000 min
Response: 7811426
Conc: 35.87 ng/ml




#26 AR-1254-1

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 17571892
Conc: 41.34 ng/ml



#26 AR-1254-1


R.T.: 5.548 min
Delta R.T.: -0.001 min
Response: 10906692
Conc: 33.77 ng/ml



#27 AR-1254-2


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 9607870
Conc: 25.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#27 AR-1254-2

R.T.: 5.696 min
Delta R.T.: -0.001 min
Response: 4262534
Conc: 15.26 ng/ml




#28 AR-1254-3

R.T.: 6.125 min
Delta R.T.: -0.001 min
Response: 15868688
Conc: 27.77 ng/ml



#28 AR-1254-3


R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 6323016
Conc: 15.09 ng/ml



#29 AR-1254-4


R.T.: 6.349 min
Delta R.T.: -0.007 min
Response: 11535842
Conc: 33.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#29 AR-1254-4

R.T.: 6.321 min
Delta R.T.: -0.005 min
Response: 2651152
Conc: 11.64 ng/ml




#30 AR-1254-5

R.T.: 6.768 min
Delta R.T.: -0.007 min
Response: 5774076
Conc: 11.27 ng/ml



#30 AR-1254-5


R.T.: 6.741 min
Delta R.T.: -0.001 min
Response: 943028
Conc: 2.90 ng/ml



#31 AR-1260-1


R.T.: 6.268 min
Delta R.T.: 0.011 min
Response: 18265131
Conc: 51.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#31 AR-1260-1

R.T.: 6.242 min
Delta R.T.: 0.013 min
Response: 3232508
Conc: 12.88 ng/ml




#32 AR-1260-2

R.T.: 6.444 min
Delta R.T.: -0.002 min
Response: 5225914
Conc: 11.06 ng/ml



#32 AR-1260-2


R.T.: 6.414 min
Delta R.T.: -0.003 min
Response: 913976
Conc: 3.11 ng/ml



#33 AR-1260-3


R.T.: 6.811 min
Delta R.T.: -0.002 min
Response: 10975960
Conc: 25.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#33 AR-1260-3

R.T.: 6.565 min
Delta R.T.: -0.004 min
Response: 624549
Conc: 2.34 ng/ml




#34 AR-1260-4

R.T.: 7.066 min
Delta R.T.: -0.006 min
Response: 2690924
Conc: 8.56 ng/ml



#34 AR-1260-4


R.T.: 7.036 min
Delta R.T.: -0.003 min
Response: 562199
Conc: 2.96 ng/ml



#35 AR-1260-5


R.T.: 7.295 min
Delta R.T.: -0.021 min
Response: 8661041
Conc: 10.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#35 AR-1260-5

R.T.: 7.282 min
Delta R.T.: 0.001 min
Response: 844106
Conc: 1.99 ng/ml




#36 AR-1262-1

R.T.: 6.811 min
Delta R.T.: -0.004 min
Response: 10975960
Conc: 18.00 ng/ml



#36 AR-1262-1


R.T.: 6.777 min
Delta R.T.: -0.005 min
Response: 1243260
Conc: 3.68 ng/ml



#37 AR-1262-2


R.T.: 7.295 min
Delta R.T.: -0.022 min
Response: 8661041
Conc: 9.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#37 AR-1262-2

R.T.: 7.282 min
Delta R.T.: 0.000 min
Response: 844106
Conc: 1.81 ng/ml




#38 AR-1262-3

R.T.: 7.613 min
Delta R.T.: 0.012 min
Response: -2226828
Conc: N.D.



#38 AR-1262-3


R.T.: 7.558 min
Delta R.T.: -0.006 min
Response: 310795
Conc: 1.87 ng/ml



#39 AR-1262-4


R.T.: 7.657 min
Delta R.T.: -0.007 min
Response: 409098
Conc: 0.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#39 AR-1262-4

R.T.: 7.624 min
Delta R.T.: -0.004 min
Response: 365923
Conc: 1.32 ng/ml




#40 AR-1262-5

R.T.: 8.177 min
Delta R.T.: 0.016 min
Response: 215047
Conc: 0.71 ng/ml



#40 AR-1262-5


R.T.: 8.118 min
Delta R.T.: -0.002 min
Response: 67570
Conc: 0.66 ng/ml



#41 AR-1268-1


R.T.: 7.613 min
Delta R.T.: 0.012 min
Response: -2226828
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#41 AR-1268-1

R.T.: 7.558 min
Delta R.T.: -0.006 min
Response: 310795
Conc: 0.65 ng/ml




#42 AR-1268-2

R.T.: 7.657 min
Delta R.T.: -0.009 min
Response: 409098
Conc: 0.42 ng/ml



#42 AR-1268-2


R.T.: 7.624 min
Delta R.T.: -0.005 min
Response: 365923
Conc: 0.88 ng/ml



#43 AR-1268-3


R.T.: 7.850 min
Delta R.T.: -0.022 min
Response: 476089
Conc: 0.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025




#43 AR-1268-3

R.T.: 7.817 min
Delta R.T.: -0.017 min
Response: 143675
Conc: 0.45 ng/ml




#44 AR-1268-4

R.T.: 8.177 min
Delta R.T.: 0.016 min
Response: 215047
Conc: 0.63 ng/ml



#44 AR-1268-4


R.T.: 8.118 min
Delta R.T.: -0.002 min
Response: 67570
Conc: 0.58 ng/ml



#45 AR-1268-5

R.T.: 8.450 min
Delta R.T.: -0.005 min
Response: 520140
Conc: 0.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2025



#45 AR-1268-5

R.T.: 8.412 min
Delta R.T.: 0.000 min
Response: 308023
Conc: 0.43 ng/ml