

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062222\
 Data File : P0087270.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2022 17:29
 Operator : YP/AJ
 Sample : N3428-01MSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CL-01-062122MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/23/2022
 Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 18:06:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.639	49120317	24812293	21.018	22.037
2)	SA Decachlor...	10.349	8.644	38559665	15207099	20.953	19.341
Target Compounds							
3)	L1 AR-1016-1	5.674	4.727	38144315	15820087	455.711	469.837
4)	L1 AR-1016-2	5.697	4.746	53874735	21773090	458.167	468.071
5)	L1 AR-1016-3	5.760	4.921	35163152	11733341	467.143	458.006
6)	L1 AR-1016-4	5.860	4.964	32527866	9213146	542.868	451.923
7)	L1 AR-1016-5	6.156	5.176	27614764	11283009	469.858m	441.499
31)	L7 AR-1260-1	7.292	6.209	41934691	20488723	416.128	437.411
32)	L7 AR-1260-2	7.550	6.396	51383319	24561681	435.992	438.493
33)	L7 AR-1260-3	7.912	6.549	32120097	22477318	378.894	422.404
34)	L7 AR-1260-4	8.141	7.021	40939314	16147063	401.979	400.657
35)	L7 AR-1260-5	8.469	7.262	73659505	38175716	405.286	406.967

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062222\
Data File : PO087270.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2022 17:29
Operator : YP/AJ
Sample : N3428-01MSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-01-062122MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 18:06:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

