

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062222\
 Data File : P0087259.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2022 14:29
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 14:53:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.641	108.4E6	53705822	46.403	47.699
2) SA Decachlor...	10.353	8.646	83846279	34486743	45.560	43.862
Target Compounds						
16) L4 AR-1242-1	5.678	4.728	28277320	12665782	439.949	478.535
17) L4 AR-1242-2	5.701	4.747	38830605	17354168	424.965	481.782
18) L4 AR-1242-3	5.764	4.922	25075909	9564127	427.217	478.617
19) L4 AR-1242-4	5.864	5.007	20172153	9229381	432.994	472.185
20) L4 AR-1242-5	6.606	5.529	21179230	11351894	446.363	486.261



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062222\
Data File : PO087259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2022 14:29
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 14:53:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.496 min
Delta R.T.: 0.003 min
Response: 108445094
Conc: 46.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 53705822
Conc: 47.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.353 min
Delta R.T.: -0.003 min
Response: 83846279
Conc: 45.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.646 min
Delta R.T.: -0.011 min
Response: 34486743
Conc: 43.86 ng/ml



#16 AR-1242-1


R.T.: 5.678 min
Delta R.T.: 0.000 min
Response: 28277320
Conc: 439.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: -0.004 min
Response: 12665782
Conc: 478.53 ng/ml




#17 AR-1242-2

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 38830605
Conc: 424.97 ng/ml



#17 AR-1242-2


R.T.: 4.747 min
Delta R.T.: -0.003 min
Response: 17354168
Conc: 481.78 ng/ml



#18 AR-1242-3


R.T.: 5.764 min
Delta R.T.: 0.002 min
Response: 25075909
Conc: 427.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.922 min
Delta R.T.: -0.005 min
Response: 9564127
Conc: 478.62 ng/ml




#19 AR-1242-4

R.T.: 5.864 min
Delta R.T.: 0.002 min
Response: 20172153
Conc: 432.99 ng/ml



#19 AR-1242-4


R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 9229381
Conc: 472.19 ng/ml



#20 AR-1242-5

R.T.: 6.606 min
Delta R.T.: 0.001 min
Response: 21179230
Conc: 446.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.529 min
Delta R.T.: -0.005 min
Response: 11351894
Conc: 486.26 ng/ml