

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062222\
 Data File : P0087303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2022 05:17
 Operator : YP/AJ
 Sample : N3449-03 10X
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC226642-2-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/23/2022
 Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 07:50:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.639	2922238	1881350	1.250	1.671 #
2) SA Decachlor...	10.349	8.642	2912475	1410520	1.583	1.794
Target Compounds						
26) L6 AR-1254-1	6.541	5.527	544402	271759	6.244	5.615
27) L6 AR-1254-2	6.757	5.674	859452	348397	6.581	8.168
28) L6 AR-1254-3	7.126	6.078	1110462	255421	8.511	3.737m#
29) L6 AR-1254-4	7.418	6.307	391473	220680	4.097	5.231m#
30) L6 AR-1254-5	7.836	6.724	534177	339938	5.113	5.598

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062222\
Data File : PO087303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2022 05:17
Operator : YP/AJ
Sample : N3449-03 10X
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC226642-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 07:50:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.492 min
Delta R.T.: -0.001 min
Response: 2922238
Conc: 1.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC226642-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 1881350
Conc: 1.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: -0.007 min
Response: 2912475
Conc: 1.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: -0.016 min
Response: 1410520
Conc: 1.79 ng/ml




#26 AR-1254-1

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 544402
Conc: 6.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC226642-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022




#26 AR-1254-1

R.T.: 5.527 min
Delta R.T.: -0.008 min
Response: 271759
Conc: 5.62 ng/ml




#27 AR-1254-2

R.T.: 6.757 min
Delta R.T.: -0.004 min
Response: 859452
Conc: 6.58 ng/ml



#27 AR-1254-2

R.T.: 5.674 min
Delta R.T.: -0.008 min
Response: 348397
Conc: 8.17 ng/ml




#28 AR-1254-3

R.T.: 7.126 min
Delta R.T.: -0.005 min
Response: 1110462
Conc: 8.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC226642-2-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022




#28 AR-1254-3

R.T.: 6.078 min
Delta R.T.: -0.007 min
Response: 255421
Conc: 3.74 ng/ml m




#29 AR-1254-4

R.T.: 7.418 min
Delta R.T.: 0.002 min
Response: 391473
Conc: 4.10 ng/ml



#29 AR-1254-4

R.T.: 6.307 min
Delta R.T.: -0.007 min
Response: 220680
Conc: 5.23 ng/ml m




#30 AR-1254-5

R.T.: 7.836 min
Delta R.T.: -0.003 min
Response: 534177
Conc: 5.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC226642-2-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/23/2022
Supervised By :Ankita Jodhani 06/23/2022



#30 AR-1254-5

R.T.: 6.724 min
Delta R.T.: -0.007 min
Response: 339938
Conc: 5.60 ng/ml