

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062222\
 Data File : P0087308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2022 08:03
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 10:41:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.638	114.1E6	56578111	48.804	50.250
2) SA Decachlor...	10.348	8.642	91099920	38191730	49.502	48.575
Target Compounds						
26) L6 AR-1254-1	6.538	5.527	42556013	25663809	488.103	530.266
27) L6 AR-1254-2	6.758	5.674	61638021	23195422	471.982	543.835
28) L6 AR-1254-3	7.127	6.076	64234679	36182029	492.328	529.427
29) L6 AR-1254-4	7.413	6.305	46693057	22693970	488.707	537.924
30) L6 AR-1254-5	7.836	6.722	50516498	32875843	483.522	541.346



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062222\
Data File : PO087308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2022 08:03
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 10:41:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.002 min
Response: 114055649
Conc: 48.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: -0.003 min
Response: 56578111
Conc: 50.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: -0.007 min
Response: 91099920
Conc: 49.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: -0.015 min
Response: 38191730
Conc: 48.57 ng/ml



#26 AR-1254-1


R.T.: 6.538 min
Delta R.T.: -0.003 min
Response: 42556013
Conc: 488.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.527 min
Delta R.T.: -0.008 min
Response: 25663809
Conc: 530.27 ng/ml




#27 AR-1254-2

R.T.: 6.758 min
Delta R.T.: -0.003 min
Response: 61638021
Conc: 471.98 ng/ml



#27 AR-1254-2


R.T.: 5.674 min
Delta R.T.: -0.008 min
Response: 23195422
Conc: 543.84 ng/ml



#28 AR-1254-3


R.T.: 7.127 min
Delta R.T.: -0.004 min
Response: 64234679
Conc: 492.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.076 min
Delta R.T.: -0.009 min
Response: 36182029
Conc: 529.43 ng/ml




#29 AR-1254-4

R.T.: 7.413 min
Delta R.T.: -0.002 min
Response: 46693057
Conc: 488.71 ng/ml



#29 AR-1254-4


R.T.: 6.305 min
Delta R.T.: -0.008 min
Response: 22693970
Conc: 537.92 ng/ml



#30 AR-1254-5

R.T.: 7.836 min
Delta R.T.: -0.004 min
Response: 50516498
Conc: 483.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.722 min
Delta R.T.: -0.009 min
Response: 32875843
Conc: 541.35 ng/ml