

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062223\
 Data File : P0095889.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2023 22:23
 Operator : YP/AJ
 Sample : 03337-04
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WC-17

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 04:14:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.634	65593406	28445484	22.544	20.959
2) SA Decachlor...	10.231	8.668	43247772	22946385	20.875	18.713

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062223\
Data File : P0095889.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2023 22:23
Operator : YP/AJ
Sample : 03337-04
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
WC-17

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 04:14:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: 0.000 min
Response: 65593406
Conc: 22.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
WC-17




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: -0.002 min
Response: 28445484
Conc: 20.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.231 min
Delta R.T.: 0.003 min
Response: 43247772
Conc: 20.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.668 min
Delta R.T.: -0.002 min
Response: 22946385
Conc: 18.71 ng/ml