

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062322\
 Data File : P0087327.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2022 15:47
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 16:35:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.639	109.7E6	55039717	46.954	48.884
2) SA Decachlor...	10.348	8.642	88268543	37154694	47.963	47.256
Target Compounds						
16) L4 AR-1242-1	5.674	4.726	30820026	13215458	479.509	499.302
17) L4 AR-1242-2	5.698	4.745	42734691	17781980	467.692	493.659
18) L4 AR-1242-3	5.761	4.920	27667520	10078312	471.370	504.348
19) L4 AR-1242-4	5.860	5.005	21997792	9831509	472.181	502.991
20) L4 AR-1242-5	6.604	5.527	22735649	12035387	479.165	515.538



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062322\
Data File : PO087327.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2022 15:47
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 16:35:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 109731619
Conc: 46.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 55039717
Conc: 48.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: -0.007 min
Response: 88268543
Conc: 47.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.642 min
Delta R.T.: -0.015 min
Response: 37154694
Conc: 47.26 ng/ml



#16 AR-1242-1


R.T.: 5.674 min
Delta R.T.: -0.002 min
Response: 30820026
Conc: 479.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.726 min
Delta R.T.: -0.006 min
Response: 13215458
Conc: 499.30 ng/ml




#17 AR-1242-2

R.T.: 5.698 min
Delta R.T.: -0.002 min
Response: 42734691
Conc: 467.69 ng/ml



#17 AR-1242-2


R.T.: 4.745 min
Delta R.T.: -0.006 min
Response: 17781980
Conc: 493.66 ng/ml



#18 AR-1242-3


R.T.: 5.761 min
Delta R.T.: -0.001 min
Response: 27667520
Conc: 471.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.920 min
Delta R.T.: -0.006 min
Response: 10078312
Conc: 504.35 ng/ml




#19 AR-1242-4

R.T.: 5.860 min
Delta R.T.: -0.001 min
Response: 21997792
Conc: 472.18 ng/ml



#19 AR-1242-4


R.T.: 5.005 min
Delta R.T.: -0.007 min
Response: 9831509
Conc: 502.99 ng/ml



#20 AR-1242-5

R.T.: 6.604 min
Delta R.T.: -0.001 min
Response: 22735649
Conc: 479.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.527 min
Delta R.T.: -0.007 min
Response: 12035387
Conc: 515.54 ng/ml