

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062323\
 Data File : P0095925.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jun 2023 17:16
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 22:39:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.634	152.5E6	69222152	52.421	51.004
2) SA Decachlor...	10.239	8.670	109.5E6	59573377	52.852	48.584
Target Compounds						
21) L5 AR-1248-1	5.596	4.724	26432247	13722781	522.165	505.973
22) L5 AR-1248-2	5.872	4.962	41672718	20145514	506.127	503.805
23) L5 AR-1248-3	6.077	5.004	43872412	21718661	509.711	504.777
24) L5 AR-1248-4	6.483	5.176	42301132	26126597	508.966	502.311
25) L5 AR-1248-5	6.522	5.571	41607891	24793942	494.985	494.225



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062323\
Data File : P0095925.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jun 2023 17:16
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 22:39:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.001 min
Response: 152520777
Conc: 52.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: 0.000 min
Response: 69222152
Conc: 51.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.011 min
Response: 109493967
Conc: 52.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.670 min
Delta R.T.: 0.000 min
Response: 59573377
Conc: 48.58 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: 0.004 min
Response: 26432247
Conc: 522.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 13722781
Conc: 505.97 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: 0.004 min
Response: 41672718
Conc: 506.13 ng/ml



#22 AR-1248-2


R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 20145514
Conc: 503.80 ng/ml



#23 AR-1248-3


R.T.: 6.077 min
Delta R.T.: 0.004 min
Response: 43872412
Conc: 509.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.004 min
Delta R.T.: 0.000 min
Response: 21718661
Conc: 504.78 ng/ml




#24 AR-1248-4

R.T.: 6.483 min
Delta R.T.: 0.004 min
Response: 42301132
Conc: 508.97 ng/ml



#24 AR-1248-4


R.T.: 5.176 min
Delta R.T.: 0.000 min
Response: 26126597
Conc: 502.31 ng/ml



#25 AR-1248-5

R.T.: 6.522 min
Delta R.T.: 0.004 min
Response: 41607891
Conc: 494.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 24793942
Conc: 494.22 ng/ml