

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062323\
 Data File : P0095955.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2023 03:53
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 06:26:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.416	3.633	161.7E6	73217335	55.563	53.948
2)	SA Decachlor...	10.239	8.669	116.8E6	60430484	56.382	49.283
Target Compounds							
26)	L6 AR-1254-1	6.458	5.528	50107708	38387631	526.254	518.305
27)	L6 AR-1254-2	6.678	5.677	73177500	34922465	526.911	511.745
28)	L6 AR-1254-3	7.046	6.082	72000392	51965983	532.936	521.852
29)	L6 AR-1254-4	7.333	6.312	48460777	31058320	540.451	526.234
30)	L6 AR-1254-5	7.756	6.732	57918196	46780649	563.038	524.637



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062323\
Data File : P0095955.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2023 03:53
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 06:26:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 161662362
Conc: 55.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: -0.002 min
Response: 73217335
Conc: 53.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.011 min
Response: 116807031
Conc: 56.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.669 min
Delta R.T.: -0.001 min
Response: 60430484
Conc: 49.28 ng/ml



#26 AR-1254-1


R.T.: 6.458 min
Delta R.T.: 0.004 min
Response: 50107708
Conc: 526.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: -0.002 min
Response: 38387631
Conc: 518.31 ng/ml




#27 AR-1254-2

R.T.: 6.678 min
Delta R.T.: 0.004 min
Response: 73177500
Conc: 526.91 ng/ml



#27 AR-1254-2


R.T.: 5.677 min
Delta R.T.: -0.002 min
Response: 34922465
Conc: 511.74 ng/ml



#28 AR-1254-3


R.T.: 7.046 min
Delta R.T.: 0.003 min
Response: 7200392
Conc: 532.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.082 min
Delta R.T.: -0.002 min
Response: 51965983
Conc: 521.85 ng/ml




#29 AR-1254-4

R.T.: 7.333 min
Delta R.T.: 0.004 min
Response: 48460777
Conc: 540.45 ng/ml



#29 AR-1254-4


R.T.: 6.312 min
Delta R.T.: -0.001 min
Response: 31058320
Conc: 526.23 ng/ml



#30 AR-1254-5

R.T.: 7.756 min
Delta R.T.: 0.004 min
Response: 57918196
Conc: 563.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.732 min
Delta R.T.: -0.002 min
Response: 46780649
Conc: 524.64 ng/ml