

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062421\
 Data File : P0079114.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Jun 2021 15:11
 Operator : DD\AJ
 Sample : M2820-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PIA-123

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 02:24:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jun 20 03:02:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.324	3.600	1539725	529296	19.509	17.809m
2) SA Decachlor...	9.925	8.778	1161726	515987	18.601	17.316

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062421\
Data File : P0079114.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Jun 2021 15:11
Operator : DD\AJ
Sample : M2820-06
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PIA-123

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 02:24:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jun 20 03:02:02 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.324 min
Delta R.T.: -0.006 min
Response: 1539725
Conc: 19.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
PIA-123




#1 Tetrachloro-m-xylene

R.T.: 3.600 min
Delta R.T.: -0.006 min
Response: 529296
Conc: 17.81 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.925 min
Delta R.T.: -0.010 min
Response: 1161726
Conc: 18.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.778 min
Delta R.T.: -0.009 min
Response: 515987
Conc: 17.32 ng/ml