

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062422\
 Data File : P0087379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2022 18:30
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO062422AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 18:51:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 24 18:17:14 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.465	3.624	137.4E6	48102777	51.941	53.959
2) SA Decachlor...	10.325	8.643	113.5E6	38602172	52.068	53.601
Target Compounds						
21) L5 AR-1248-1	5.642	4.708	28609665	9907933	498.426	493.301
22) L5 AR-1248-2	5.918	4.944	40958214	13424532	488.055	493.564
23) L5 AR-1248-3	6.124	4.986	44100080	13832855	487.481	493.793
24) L5 AR-1248-4	6.529	5.157	48370317	16061961	492.152	491.883
25) L5 AR-1248-5	6.568	5.550	46618985	15241148	489.065	495.023



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062422\
Data File : P0087379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2022 18:30
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO062422AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 18:51:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 24 18:17:14 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.465 min
Delta R.T.: 0.000 min
Response: 137395036
Conc: 51.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062422AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 48102777
Conc: 53.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.325 min
Delta R.T.: 0.005 min
Response: 113504048
Conc: 52.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: 0.000 min
Response: 38602172
Conc: 53.60 ng/ml



#21 AR-1248-1


R.T.: 5.642 min
Delta R.T.: -0.001 min
Response: 28609665
Conc: 498.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062422AR1248




#21 AR-1248-1

R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 9907933
Conc: 493.30 ng/ml




#22 AR-1248-2

R.T.: 5.918 min
Delta R.T.: -0.001 min
Response: 40958214
Conc: 488.05 ng/ml



#22 AR-1248-2


R.T.: 4.944 min
Delta R.T.: 0.000 min
Response: 13424532
Conc: 493.56 ng/ml



#23 AR-1248-3


R.T.: 6.124 min
Delta R.T.: -0.001 min
Response: 44100080
Conc: 487.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062422AR1248




#23 AR-1248-3

R.T.: 4.986 min
Delta R.T.: 0.000 min
Response: 13832855
Conc: 493.79 ng/ml




#24 AR-1248-4

R.T.: 6.529 min
Delta R.T.: -0.001 min
Response: 48370317
Conc: 492.15 ng/ml



#24 AR-1248-4


R.T.: 5.157 min
Delta R.T.: 0.000 min
Response: 16061961
Conc: 491.88 ng/ml



#25 AR-1248-5

R.T.: 6.568 min
Delta R.T.: -0.001 min
Response: 46618985
Conc: 489.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO062422AR1248



#25 AR-1248-5

R.T.: 5.550 min
Delta R.T.: 0.000 min
Response: 15241148
Conc: 495.02 ng/ml