

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062520\
 Data File : P0069184.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2020 9:54
 Operator : DD\AJ
 Sample : L3094-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BRENNAN-FL-001

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 13:33:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 19 05:15:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.403	3.564	816360	929585	27.595	24.152
2)	SA Decachlor...	10.068	8.771	1058430	1105005	22.285	20.299
Target Compounds							
3)	L1 AR-1016-1	5.646	0.000	36066	0	27.866	N.D. #
10)	L2 AR-1221-3	0.000	4.045	0	48812	N.D.	45.444 #
11)	L3 AR-1232-1	0.000	4.045	0	48812	N.D.	55.426 #
16)	L4 AR-1242-1	5.646	0.000	36066	0	36.257	N.D. #
20)	L4 AR-1242-5	6.684	5.675	73349	401769	82.200	361.021 #
21)	L5 AR-1248-1	5.646	0.000	36066	0	46.262	N.D. #
24)	L5 AR-1248-4	6.618	0.000	526195	0	349.871	N.D. #
25)	L5 AR-1248-5	6.684	5.675	73349	401769	50.652	277.713 #
26)	L6 AR-1254-1	6.618	5.675	526195	401769	347.333	170.886 #
28)	L6 AR-1254-3	7.230	6.247	188735	74828	71.088	21.701 #
29)	L6 AR-1254-4	7.507	6.516	14186	176517	6.652	75.696 #
32)	L7 AR-1260-2	0.000	6.516	0	176517	N.D.	62.011 #
35)	L7 AR-1260-5	8.543	0.000	35713	0	8.125	N.D. #
37)	L8 AR-1262-2	8.543	0.000	35713	0	7.436	N.D. #
38)	L8 AR-1262-3	0.000	7.690	0	341827	N.D.	137.719 #
40)	L8 AR-1262-5	9.458	0.000	12971	0	7.229	N.D. #
41)	L9 AR-1268-1	0.000	7.690	0	341827	N.D.	46.901 #
44)	L9 AR-1268-4	9.458	0.000	12971	0	6.190	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062520\
Data File : P0069184.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2020 9:54
Operator : DD\AJ
Sample : L3094-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 13:33:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 19 05:15:02 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: -0.002 min
Response: 816360
Conc: 27.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: 0.000 min
Response: 929585
Conc: 24.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: -0.002 min
Response: 1058430
Conc: 22.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.771 min
Delta R.T.: 0.000 min
Response: 1105005
Conc: 20.30 ng/ml



#3 AR-1016-1


R.T.: 5.646 min
Delta R.T.: -0.066 min
Response: 36066
Conc: 27.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.796 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.841 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.045 min
Delta R.T.: 0.046 min
Response: 48812
Conc: 45.44 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 4.842 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#11 AR-1232-1

R.T.: 4.045 min
Delta R.T.: 0.045 min
Response: 48812
Conc: 55.43 ng/ml




#16 AR-1242-1

R.T.: 5.646 min
Delta R.T.: -0.067 min
Response: 36066
Conc: 36.26 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T.: 4.796 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.684 min
Delta R.T.: 0.005 min
Response: 73349
Conc: 82.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#20 AR-1242-5

R.T.: 5.675 min
Delta R.T.: 0.021 min
Response: 401769
Conc: 361.02 ng/ml




#21 AR-1248-1

R.T.: 5.646 min
Delta R.T.: -0.066 min
Response: 36066
Conc: 46.26 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 4.795 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.618 min
Delta R.T.: -0.023 min
Response: 526195
Conc: 349.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.277 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.684 min
Delta R.T.: 0.006 min
Response: 73349
Conc: 50.65 ng/ml



#25 AR-1248-5


R.T.: 5.675 min
Delta R.T.: -0.020 min
Response: 401769
Conc: 277.71 ng/ml



#26 AR-1254-1


R.T.: 6.618 min
Delta R.T.: 0.011 min
Response: 526195
Conc: 347.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#26 AR-1254-1

R.T.: 5.675 min
Delta R.T.: 0.023 min
Response: 401769
Conc: 170.89 ng/ml




#28 AR-1254-3

R.T.: 7.230 min
Delta R.T.: 0.017 min
Response: 188735
Conc: 71.09 ng/ml



#28 AR-1254-3


R.T.: 6.247 min
Delta R.T.: 0.022 min
Response: 74828
Conc: 21.70 ng/ml



#29 AR-1254-4


R.T.: 7.507 min
Delta R.T.: -0.001 min
Response: 14186
Conc: 6.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#29 AR-1254-4

R.T.: 6.516 min
Delta R.T.: 0.051 min
Response: 176517
Conc: 75.70 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.644 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.516 min
Delta R.T.: -0.046 min
Response: 176517
Conc: 62.01 ng/ml



#35 AR-1260-5


R.T.: 8.543 min
Delta R.T.: -0.001 min
Response: 35713
Conc: 8.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.440 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 8.543 min
Delta R.T.: -0.002 min
Response: 35713
Conc: 7.44 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.441 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 8.816 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#38 AR-1262-3

R.T.: 7.690 min
Delta R.T.: -0.033 min
Response: 341827
Conc: 137.72 ng/ml




#40 AR-1262-5

R.T.: 9.458 min
Delta R.T.: -0.001 min
Response: 12971
Conc: 7.23 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.283 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.816 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
BRENNAN-FL-001




#41 AR-1268-1

R.T.: 7.690 min
Delta R.T.: -0.034 min
Response: 341827
Conc: 46.90 ng/ml



#44 AR-1268-4

R.T.: 9.458 min
Delta R.T.: -0.002 min
Response: 12971
Conc: 6.19 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 8.283 min
Response: 0
Conc: N.D.