

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062520\
 Data File : P0069199.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2020 14:42
 Operator : DD\AJ
 Sample : PB129761BL
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB129761BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 16:12:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 19 05:15:02 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.403	3.564	674794	805855	22.810	20.937
2) SA Decachlor...	10.066	8.770	833455	877148	17.548	16.113
Target Compounds						
26) L6 AR-1254-1	6.552	0.000	50170	0	33.116	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062520\
Data File : P0069199.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2020 14:42
Operator : DD\AJ
Sample : PB129761BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB129761BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 16:12:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 19 05:15:02 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: -0.002 min
Response: 674794
Conc: 22.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB129761BL




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: 0.000 min
Response: 805855
Conc: 20.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.066 min
Delta R.T.: -0.003 min
Response: 833455
Conc: 17.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.770 min
Delta R.T.: 0.000 min
Response: 877148
Conc: 16.11 ng/ml



#26 AR-1254-1

R.T.: 6.552 min
Delta R.T.: -0.055 min
Response: 50170
Conc: 33.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB129761BL



#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.652 min
Response: 0
Conc: N.D.