

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062521\
 Data File : P0079138.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2021 10:11
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 14:16:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jun 20 03:02:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.323	3.600	4380021	1579958	55.498	53.160
2) SA Decachlor...	9.922	8.777	2956169	1441750	47.333	48.383
Target Compounds						
21) L5 AR-1248-1	5.618	4.825	840005	319076	531.815	501.160
22) L5 AR-1248-2	5.908	5.084	1181178	487333	533.922	493.661
23) L5 AR-1248-3	6.119	5.126	1367336	484965	527.298	488.493
24) L5 AR-1248-4	6.542	5.307	1484082	576131	535.454	490.659
25) L5 AR-1248-5	6.579	5.722	1465825	540040	536.523	501.534



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062521\
Data File : P0079138.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2021 10:11
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 14:16:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jun 20 03:02:02 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.323 min
Delta R.T.: -0.007 min
Response: 4380021
Conc: 55.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.600 min
Delta R.T.: -0.006 min
Response: 1579958
Conc: 53.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.922 min
Delta R.T.: -0.012 min
Response: 2956169
Conc: 47.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.777 min
Delta R.T.: -0.010 min
Response: 1441750
Conc: 48.38 ng/ml



#21 AR-1248-1


R.T.: 5.618 min
Delta R.T.: -0.008 min
Response: 840005
Conc: 531.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.825 min
Delta R.T.: -0.008 min
Response: 319076
Conc: 501.16 ng/ml




#22 AR-1248-2

R.T.: 5.908 min
Delta R.T.: -0.009 min
Response: 1181178
Conc: 533.92 ng/ml



#22 AR-1248-2


R.T.: 5.084 min
Delta R.T.: -0.008 min
Response: 487333
Conc: 493.66 ng/ml



#23 AR-1248-3


R.T.: 6.119 min
Delta R.T.: -0.009 min
Response: 1367336
Conc: 527.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.126 min
Delta R.T.: -0.008 min
Response: 484965
Conc: 488.49 ng/ml




#24 AR-1248-4

R.T.: 6.542 min
Delta R.T.: -0.010 min
Response: 1484082
Conc: 535.45 ng/ml



#24 AR-1248-4


R.T.: 5.307 min
Delta R.T.: -0.008 min
Response: 576131
Conc: 490.66 ng/ml



#25 AR-1248-5

R.T.: 6.579 min
Delta R.T.: -0.009 min
Response: 1465825
Conc: 536.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.722 min
Delta R.T.: -0.008 min
Response: 540040
Conc: 501.53 ng/ml