

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062521\
 Data File : P0079139.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2021 10:28
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 14:17:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jun 20 03:02:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.323	3.600	4239709	1535848	53.720	51.676
2) SA Decachlor...	9.924	8.777	2990112	1428110	47.877	47.925
Target Compounds						
26) L6 AR-1254-1	6.511	5.679	1440550	805938	534.817	485.100
27) L6 AR-1254-2	6.739	5.837	2189069	713449	521.901	484.117
28) L6 AR-1254-3	7.114	6.250	2253645	1085745	506.029	488.147
29) L6 AR-1254-4	7.407	6.488	1590256	686549	516.825	504.933
30) L6 AR-1254-5	7.830	6.909	1739939	993856	533.015	499.134



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062521\
Data File : P0079139.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2021 10:28
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 14:17:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jun 20 03:02:02 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.323 min
Delta R.T.: -0.007 min
Response: 4239709
Conc: 53.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.600 min
Delta R.T.: -0.006 min
Response: 1535848
Conc: 51.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.924 min
Delta R.T.: -0.011 min
Response: 2990112
Conc: 47.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.777 min
Delta R.T.: -0.010 min
Response: 1428110
Conc: 47.93 ng/ml



#26 AR-1254-1


R.T.: 6.511 min
Delta R.T.: -0.008 min
Response: 1440550
Conc: 534.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.679 min
Delta R.T.: -0.007 min
Response: 805938
Conc: 485.10 ng/ml




#27 AR-1254-2

R.T.: 6.739 min
Delta R.T.: -0.008 min
Response: 2189069
Conc: 521.90 ng/ml



#27 AR-1254-2


R.T.: 5.837 min
Delta R.T.: -0.008 min
Response: 713449
Conc: 484.12 ng/ml



#28 AR-1254-3


R.T.: 7.114 min
Delta R.T.: -0.008 min
Response: 2253645
Conc: 506.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.250 min
Delta R.T.: -0.008 min
Response: 1085745
Conc: 488.15 ng/ml




#29 AR-1254-4

R.T.: 7.407 min
Delta R.T.: -0.008 min
Response: 1590256
Conc: 516.83 ng/ml



#29 AR-1254-4


R.T.: 6.488 min
Delta R.T.: -0.008 min
Response: 686549
Conc: 504.93 ng/ml



#30 AR-1254-5

R.T.: 7.830 min
Delta R.T.: -0.008 min
Response: 1739939
Conc: 533.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.909 min
Delta R.T.: -0.009 min
Response: 993856
Conc: 499.13 ng/ml