

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062525\
 Data File : P0111839.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2025 10:49
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 13:28:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.676	3.671	322.0E6	296.9E6	55.910	52.861
2) SA Decachlor...	8.709	8.657	281.8E6	93521414	53.679	52.609
Target Compounds						
21) L5 AR-1248-1	4.764	4.748	81459953	69670869	505.129	522.723
22) L5 AR-1248-2	5.002	4.984	109.6E6	93975422	508.461	527.491
23) L5 AR-1248-3	5.216	5.026	140.3E6	98528740	519.968	520.111
24) L5 AR-1248-4	5.570	5.196	206.7E6	117.1E6	517.018	512.967
25) L5 AR-1248-5	5.611	5.587	144.6E6	115.2E6	515.022	529.072



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062525\
Data File : P0111839.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2025 10:49
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 13:28:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.676 min
Delta R.T.: 0.000 min
Response: 322008163
Conc: 55.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.671 min
Delta R.T.: -0.002 min
Response: 296936866
Conc: 52.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.709 min
Delta R.T.: -0.004 min
Response: 281773397
Conc: 53.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.657 min
Delta R.T.: -0.004 min
Response: 93521414
Conc: 52.61 ng/ml



#21 AR-1248-1


R.T.: 4.764 min
Delta R.T.: 0.000 min
Response: 81459953
Conc: 505.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.748 min
Delta R.T.: -0.002 min
Response: 69670869
Conc: 522.72 ng/ml




#22 AR-1248-2

R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 109626918
Conc: 508.46 ng/ml



#22 AR-1248-2


R.T.: 4.984 min
Delta R.T.: -0.002 min
Response: 93975422
Conc: 527.49 ng/ml



#23 AR-1248-3


R.T.: 5.216 min
Delta R.T.: 0.000 min
Response: 140255568
Conc: 519.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.026 min
Delta R.T.: -0.002 min
Response: 98528740
Conc: 520.11 ng/ml




#24 AR-1248-4

R.T.: 5.570 min
Delta R.T.: 0.000 min
Response: 206705378
Conc: 517.02 ng/ml



#24 AR-1248-4


R.T.: 5.196 min
Delta R.T.: -0.002 min
Response: 117122473
Conc: 512.97 ng/ml



#25 AR-1248-5

R.T.: 5.611 min
Delta R.T.: 0.000 min
Response: 144623665
Conc: 515.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.587 min
Delta R.T.: -0.002 min
Response: 115213223
Conc: 529.07 ng/ml