

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0062618\
Data File : P0046278.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jun 2018 19:06
Operator : SM/UA
Sample : PB110592BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB110592BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 02:18:33 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 26 05:14:56 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.380	3.607	12282763	5618007	21.015	22.634
2) SA Decachlor...	10.044	8.572	6991649	3116404	15.164	18.591

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0062618\
Data File : P0046278.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jun 2018 19:06
Operator : SM/UA
Sample : PB110592BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB110592BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 02:18:33 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 26 05:14:56 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.380 min
Delta R.T.: 0.002 min
Response: 12282763
Conc: 21.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB110592BL




#1 Tetrachloro-m-xylene

R.T.: 3.607 min
Delta R.T.: 0.002 min
Response: 5618007
Conc: 22.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.044 min
Delta R.T.: -0.002 min
Response: 6991649
Conc: 15.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.572 min
Delta R.T.: -0.005 min
Response: 3116404
Conc: 18.59 ng/ml