

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062623\
 Data File : P0095975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jun 2023 14:15
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 22:02:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.420	3.637	150.1E6	71250348	51.602	52.499
2) SA Decachlor...	10.251	8.676	112.6E6	60770445	54.358	49.560
Target Compounds						
21) L5 AR-1248-1	5.600	4.727	26309293	14390506	519.736	530.593
22) L5 AR-1248-2	5.877	4.966	40681604	21436701	494.090	536.095
23) L5 AR-1248-3	6.082	5.007	42811624	22441960	497.387	521.588
24) L5 AR-1248-4	6.489	5.179	41756946	26191938	502.418	503.567
25) L5 AR-1248-5	6.528	5.574	40424748	24480780	480.910	487.983



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062623\
Data File : PO095975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jun 2023 14:15
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 22:02:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.420 min
Delta R.T.: 0.005 min
Response: 150137877
Conc: 51.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: 0.002 min
Response: 71250348
Conc: 52.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.251 min
Delta R.T.: 0.023 min
Response: 112613831
Conc: 54.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.006 min
Response: 60770445
Conc: 49.56 ng/ml



#21 AR-1248-1


R.T.: 5.600 min
Delta R.T.: 0.009 min
Response: 26309293
Conc: 519.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.727 min
Delta R.T.: 0.004 min
Response: 14390506
Conc: 530.59 ng/ml




#22 AR-1248-2

R.T.: 5.877 min
Delta R.T.: 0.008 min
Response: 40681604
Conc: 494.09 ng/ml



#22 AR-1248-2


R.T.: 4.966 min
Delta R.T.: 0.003 min
Response: 21436701
Conc: 536.09 ng/ml



#23 AR-1248-3


R.T.: 6.082 min
Delta R.T.: 0.009 min
Response: 42811624
Conc: 497.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.007 min
Delta R.T.: 0.004 min
Response: 22441960
Conc: 521.59 ng/ml




#24 AR-1248-4

R.T.: 6.489 min
Delta R.T.: 0.010 min
Response: 41756946
Conc: 502.42 ng/ml



#24 AR-1248-4


R.T.: 5.179 min
Delta R.T.: 0.004 min
Response: 26191938
Conc: 503.57 ng/ml



#25 AR-1248-5

R.T.: 6.528 min
Delta R.T.: 0.010 min
Response: 40424748
Conc: 480.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.574 min
Delta R.T.: 0.004 min
Response: 24480780
Conc: 487.98 ng/ml