

CASE NARRATIVE

Aqua Survey, Inc.

Project Name: 41-062-GBA-Debbies Creek Channel

Project # N/A

Chemtech Project # M2771

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 06/17/2021.

2 Water samples were received on 06/17/2021.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, Dissolved ICP-TAL Metals, Dissolved Mercury, DISSOLVED METALS-TAL, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, PESTICIDE Group1, SVOCMS Group1, Trivalent Chromium and TSS. This data package contains results for SVOCMS Group1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_G using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {M2777-01MS} with File ID: BG049032.D recoveries met the requirements for all compounds except for Acetophenone[232%], Benzidine[97%], Hexachloroethane[211%] and N-Nitrosodiphenylamine[174%] due to matrix interference.

The MSD {M2777-01MSD} with File ID: BG049033.D recoveries met the acceptable requirements except for Acetophenone[232%], Atrazine[147%], Benzidine[97%], Hexachloroethane[216%] and N-Nitrosodiphenylamine[174%] due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

The %RSD is greater than 20% for certain compounds in the Initial Calibration (Method 8270E-BG061121.M) Compound 2-Nitroaniline is passing on Quadratic Regression.

The %RSD is greater than 15% for certain compounds in the Initial Calibration (Method 8270E-BP060821.M) Compound Hexachlorocyclopentadiene, 2,4-Dinitrophenol are passing on Linear Regression.

The Continuous Calibration File ID BG049028.D met the requirements except for 2,4-Dinitrotoluene The associate samples have no positive hit for these compounds therefore no corrective action was required .

The Continuous Calibration File ID BG049038.D met the requirements except for 2,2-oxybis(1-Chloropropane),2,4-Dinitrophenol,4,6-Dinitro-2-methylphenol and Benzidine The associate samples have no positive hit for these compounds therefore no corrective action was required.

The Continuous Calibration File ID BP006008.D met the requirements except for Benzaldehyde and Caprolactam The associate samples have no positive hit for these compounds therefore no corrective action was required.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____