

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062822\
 Data File : P0087521.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2022 22:52
 Operator : YP/AJ
 Sample : N3491-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AUD-1728

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/29/2022
 Supervised By :Ankita Jodhani 06/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 02:53:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.451	3.612	12820165	16907748	4.953m	17.183m#
2) SA Decachlor...	10.304	8.617	2065995	602762	1.011	0.744m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO062822\
Data File : PO087521.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2022 22:52
Operator : YP/AJ
Sample : N3491-01
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AUD-1728


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/29/2022
Supervised By :Ankita Jodhani 06/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 02:53:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.451 min
Delta R.T.: 0.000 min
Response: 12820165
Conc: 4.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AUD-1728


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/29/2022
Supervised By :Ankita Jodhani 06/29/2022




#1 Tetrachloro-m-xylene

R.T.: 3.612 min
Delta R.T.: 0.000 min
Response: 16907748
Conc: 17.18 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.304 min
Delta R.T.: -0.001 min
Response: 2065995
Conc: 1.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.617 min
Delta R.T.: 0.002 min
Response: 602762
Conc: 0.74 ng/ml m