

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062824\
 Data File : P0104551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2024 18:16
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 23:50:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 21 04:29:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.405	3.546	411.5E6	66216145	51.802	44.118
2) SA Decachlor...	10.114	8.468	292.8E6	159.2E6	54.494	49.614
Target Compounds						
26) L6 AR-1254-1	6.418	5.392	139.2E6	27479098	467.623	393.373
27) L6 AR-1254-2	6.636	5.538	213.9E6	24297332	483.211	389.057
28) L6 AR-1254-3	7.002	5.936	218.6E6	48173243	484.933	401.627
29) L6 AR-1254-4	7.286	6.162	160.0E6	34998305	480.068	436.164
30) L6 AR-1254-5	7.704	6.576	160.8E6	51825963	464.833	444.587



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062824\
Data File : PO104551.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2024 18:16
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 23:50:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062024.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 21 04:29:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.001 min
Response: 411497994
Conc: 51.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: -0.001 min
Response: 66216145
Conc: 44.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.114 min
Delta R.T.: 0.011 min
Response: 292786871
Conc: 54.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.468 min
Delta R.T.: 0.005 min
Response: 159231851
Conc: 49.61 ng/ml



#26 AR-1254-1


R.T.: 6.418 min
Delta R.T.: 0.001 min
Response: 139208847
Conc: 467.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.392 min
Delta R.T.: -0.002 min
Response: 27479098
Conc: 393.37 ng/ml




#27 AR-1254-2

R.T.: 6.636 min
Delta R.T.: 0.002 min
Response: 213920489
Conc: 483.21 ng/ml



#27 AR-1254-2


R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 24297332
Conc: 389.06 ng/ml



#28 AR-1254-3


R.T.: 7.002 min
Delta R.T.: 0.002 min
Response: 218608337
Conc: 484.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.936 min
Delta R.T.: -0.001 min
Response: 48173243
Conc: 401.63 ng/ml




#29 AR-1254-4

R.T.: 7.286 min
Delta R.T.: 0.002 min
Response: 159955213
Conc: 480.07 ng/ml



#29 AR-1254-4


R.T.: 6.162 min
Delta R.T.: -0.001 min
Response: 34998305
Conc: 436.16 ng/ml



#30 AR-1254-5

R.T.: 7.704 min
Delta R.T.: 0.002 min
Response: 160776723
Conc: 464.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.576 min
Delta R.T.: 0.000 min
Response: 51825963
Conc: 444.59 ng/ml