

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062922\
 Data File : P0087547.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 09:42
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 11:59:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.607	135.1E6	50369393	52.188	51.189
2) SA Decachlor...	10.300	8.610	108.5E6	34913502	53.130	43.068
Target Compounds						
16) L4 AR-1242-1	5.625	4.687	36911203	14798499	509.028	518.512
17) L4 AR-1242-2	5.648	4.706	50757642	20285051	497.121	520.003
18) L4 AR-1242-3	5.710	4.881	32325252	11020008	503.323	528.691
19) L4 AR-1242-4	5.810	4.964	25797398	11111098	506.742	525.923
20) L4 AR-1242-5	6.551	5.488	27162138	14178694	511.593	542.415



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062922\
Data File : PO087547.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2022 09:42
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 11:59:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.002 min
Response: 135079682
Conc: 52.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.607 min
Delta R.T.: -0.004 min
Response: 50369393
Conc: 51.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.300 min
Delta R.T.: -0.005 min
Response: 108526372
Conc: 53.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.610 min
Delta R.T.: -0.005 min
Response: 34913502
Conc: 43.07 ng/ml



#16 AR-1242-1


R.T.: 5.625 min
Delta R.T.: -0.003 min
Response: 36911203
Conc: 509.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.687 min
Delta R.T.: -0.004 min
Response: 14798499
Conc: 518.51 ng/ml




#17 AR-1242-2

R.T.: 5.648 min
Delta R.T.: -0.003 min
Response: 50757642
Conc: 497.12 ng/ml



#17 AR-1242-2


R.T.: 4.706 min
Delta R.T.: -0.004 min
Response: 20285051
Conc: 520.00 ng/ml



#18 AR-1242-3


R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 32325252
Conc: 503.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.881 min
Delta R.T.: -0.003 min
Response: 11020008
Conc: 528.69 ng/ml




#19 AR-1242-4

R.T.: 5.810 min
Delta R.T.: -0.004 min
Response: 25797398
Conc: 506.74 ng/ml



#19 AR-1242-4


R.T.: 4.964 min
Delta R.T.: -0.004 min
Response: 11111098
Conc: 525.92 ng/ml



#20 AR-1242-5

R.T.: 6.551 min
Delta R.T.: -0.004 min
Response: 27162138
Conc: 511.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.488 min
Delta R.T.: -0.003 min
Response: 14178694
Conc: 542.42 ng/ml