

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062923\
 Data File : P0096077.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2023 18:32
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 20:56:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.419	3.635	136.6E6	62772953	46.943	46.252
2)	SA Decachlor...	10.246	8.670	103.4E6	53992829	49.924	44.032
Target Compounds							
16)	L4 AR-1242-1	5.599	4.724	30247150	16118942	457.596	445.863
17)	L4 AR-1242-2	5.621	4.743	43303638	22320034	454.159	454.983
18)	L4 AR-1242-3	5.684	4.920	28224902	12199283	444.867	454.946
19)	L4 AR-1242-4	5.783	5.004	21802874	13164382	445.542	445.269
20)	L4 AR-1242-5	6.526	5.529	19615932	16108750	396.716	453.763



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062923\
Data File : PO096077.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2023 18:32
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 20:56:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.419 min
Delta R.T.: 0.004 min
Response: 136582853
Conc: 46.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.635 min
Delta R.T.: 0.000 min
Response: 62772953
Conc: 46.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.018 min
Response: 103427882
Conc: 49.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.670 min
Delta R.T.: 0.000 min
Response: 53992829
Conc: 44.03 ng/ml



#16 AR-1242-1


R.T.: 5.599 min
Delta R.T.: 0.007 min
Response: 30247150
Conc: 457.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 16118942
Conc: 445.86 ng/ml




#17 AR-1242-2

R.T.: 5.621 min
Delta R.T.: 0.008 min
Response: 43303638
Conc: 454.16 ng/ml



#17 AR-1242-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 22320034
Conc: 454.98 ng/ml



#18 AR-1242-3


R.T.: 5.684 min
Delta R.T.: 0.008 min
Response: 28224902
Conc: 444.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.920 min
Delta R.T.: 0.000 min
Response: 12199283
Conc: 454.95 ng/ml




#19 AR-1242-4

R.T.: 5.783 min
Delta R.T.: 0.007 min
Response: 21802874
Conc: 445.54 ng/ml



#19 AR-1242-4


R.T.: 5.004 min
Delta R.T.: 0.000 min
Response: 13164382
Conc: 445.27 ng/ml



#20 AR-1242-5

R.T.: 6.526 min
Delta R.T.: 0.009 min
Response: 19615932
Conc: 396.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.529 min
Delta R.T.: 0.000 min
Response: 16108750
Conc: 453.76 ng/ml