

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062923\
 Data File : P0096087.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2023 21:21
 Operator : YP/AJ
 Sample : 03470-01 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC216351-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/30/2023
 Supervised By :Ankita Jodhani 06/30/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 04:56:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 13:08:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.634	4986459	2285628	1.714	1.684
2) SA Decachlor...	10.243	8.667	4852578	2393560	2.342	1.952
Target Compounds						
26) L6 AR-1254-1	6.459	5.528	601215	606933	6.314	8.195m#
27) L6 AR-1254-2	6.681	5.676	1710687	834818	12.318	12.233
28) L6 AR-1254-3	7.050	6.082	1889122	1116854	13.983	11.216
29) L6 AR-1254-4	7.338	6.311	1077266	637374	12.014	10.799
30) L6 AR-1254-5	7.757	6.731	1812485	980003	17.620m	10.991 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO062923\
Data File : PO096087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2023 21:21
Operator : YP/AJ
Sample : 03470-01 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC216351-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2023
Supervised By :Ankita Jodhani 06/30/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 04:56:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 13:08:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.418 min
Delta R.T.: 0.003 min
Response: 4986459
Conc: 1.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC216351-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2023
Supervised By :Ankita Jodhani 06/30/2023




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: -0.001 min
Response: 2285628
Conc: 1.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: 0.015 min
Response: 4852578
Conc: 2.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.667 min
Delta R.T.: -0.003 min
Response: 2393560
Conc: 1.95 ng/ml




#26 AR-1254-1

R.T.: 6.459 min
Delta R.T.: 0.005 min
Response: 601215
Conc: 6.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC216351-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2023
Supervised By :Ankita Jodhani 06/30/2023




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: -0.003 min
Response: 606933
Conc: 8.19 ng/ml m




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: 0.007 min
Response: 1710687
Conc: 12.32 ng/ml



#27 AR-1254-2

R.T.: 5.676 min
Delta R.T.: -0.003 min
Response: 834818
Conc: 12.23 ng/ml




#28 AR-1254-3

R.T.: 7.050 min
Delta R.T.: 0.007 min
Response: 1889122
Conc: 13.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC216351-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2023
Supervised By :Ankita Jodhani 06/30/2023




#28 AR-1254-3

R.T.: 6.082 min
Delta R.T.: -0.002 min
Response: 1116854
Conc: 11.22 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.009 min
Response: 1077266
Conc: 12.01 ng/ml



#29 AR-1254-4

R.T.: 6.311 min
Delta R.T.: -0.002 min
Response: 637374
Conc: 10.80 ng/ml




#30 AR-1254-5

R.T.: 7.757 min
Delta R.T.: 0.006 min
Response: 1812485
Conc: 17.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC216351-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2023
Supervised By :Ankita Jodhani 06/30/2023



#30 AR-1254-5

R.T.: 6.731 min
Delta R.T.: -0.003 min
Response: 980003
Conc: 10.99 ng/ml