

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070121\
 Data File : P0079276.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jul 2021 4:54
 Operator : DD\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO070121AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 07:38:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 02 07:33:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.020	4.224	3265926	1488815	56.046	50.609
2)	SA Decachlor...	11.149	9.663	2407555	1170319	57.007	44.739
Target Compounds							
21)	L5 AR-1248-1	6.351	5.506	635988	326688	509.387	478.902
22)	L5 AR-1248-2	6.647	5.774	836310	446448	510.854	473.144
23)	L5 AR-1248-3	6.866	5.820	990298	468860	516.393	476.801
24)	L5 AR-1248-4	7.297	6.008	1120400	539492	512.331	483.886
25)	L5 AR-1248-5	7.337	6.437	1089110	516085	512.434	474.574



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070121\
Data File : P0079276.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2021 4:54
Operator : DD\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO070121AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 07:38:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070121.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 02 07:33:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.020 min
Delta R.T.: 0.000 min
Response: 3265926
Conc: 56.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO070121AR1248




#1 Tetrachloro-m-xylene

R.T.: 4.224 min
Delta R.T.: -0.001 min
Response: 1488815
Conc: 50.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.149 min
Delta R.T.: 0.002 min
Response: 2407555
Conc: 57.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.663 min
Delta R.T.: 0.009 min
Response: 1170319
Conc: 44.74 ng/ml



#21 AR-1248-1


R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 635988
Conc: 509.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO070121AR1248




#21 AR-1248-1

R.T.: 5.506 min
Delta R.T.: 0.000 min
Response: 326688
Conc: 478.90 ng/ml




#22 AR-1248-2

R.T.: 6.647 min
Delta R.T.: 0.000 min
Response: 836310
Conc: 510.85 ng/ml



#22 AR-1248-2


R.T.: 5.774 min
Delta R.T.: 0.001 min
Response: 446448
Conc: 473.14 ng/ml



#23 AR-1248-3


R.T.: 6.866 min
Delta R.T.: 0.000 min
Response: 990298
Conc: 516.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO070121AR1248




#23 AR-1248-3

R.T.: 5.820 min
Delta R.T.: 0.001 min
Response: 468860
Conc: 476.80 ng/ml




#24 AR-1248-4

R.T.: 7.297 min
Delta R.T.: 0.000 min
Response: 1120400
Conc: 512.33 ng/ml



#24 AR-1248-4


R.T.: 6.008 min
Delta R.T.: 0.002 min
Response: 539492
Conc: 483.89 ng/ml



#25 AR-1248-5

R.T.: 7.337 min
Delta R.T.: 0.000 min
Response: 1089110
Conc: 512.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO070121AR1248



#25 AR-1248-5

R.T.: 6.437 min
Delta R.T.: 0.003 min
Response: 516085
Conc: 474.57 ng/ml