

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070122\
 Data File : P0087719.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2022 19:08
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 22:42:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.449	3.609	130.2E6	53931920	50.321	54.809
2)	SA Decachlor...	10.306	8.614	95415746	32807438	46.711	40.470
Target Compounds							
16)	L4 AR-1242-1	5.626	4.688	30466571	14912330	420.153	522.500
17)	L4 AR-1242-2	5.649	4.706	42329024	20738722	414.571	531.632 #
18)	L4 AR-1242-3	5.711	4.882	26282317	11205890	409.231	537.609 #
19)	L4 AR-1242-4	5.810	4.966	20284136	11283774	398.444	534.096 #
20)	L4 AR-1242-5	6.552	5.489	23761522	14382223	447.543	550.202



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070122\
Data File : PO087719.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2022 19:08
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 22:42:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.002 min
Response: 130246910
Conc: 50.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.609 min
Delta R.T.: -0.002 min
Response: 53931920
Conc: 54.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.306 min
Delta R.T.: 0.000 min
Response: 95415746
Conc: 46.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.614 min
Delta R.T.: -0.001 min
Response: 32807438
Conc: 40.47 ng/ml



#16 AR-1242-1


R.T.: 5.626 min
Delta R.T.: -0.002 min
Response: 30466571
Conc: 420.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.688 min
Delta R.T.: -0.002 min
Response: 14912330
Conc: 522.50 ng/ml




#17 AR-1242-2

R.T.: 5.649 min
Delta R.T.: -0.002 min
Response: 42329024
Conc: 414.57 ng/ml



#17 AR-1242-2


R.T.: 4.706 min
Delta R.T.: -0.003 min
Response: 20738722
Conc: 531.63 ng/ml



#18 AR-1242-3


R.T.: 5.711 min
Delta R.T.: -0.003 min
Response: 26282317
Conc: 409.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.882 min
Delta R.T.: -0.002 min
Response: 11205890
Conc: 537.61 ng/ml




#19 AR-1242-4

R.T.: 5.810 min
Delta R.T.: -0.003 min
Response: 20284136
Conc: 398.44 ng/ml



#19 AR-1242-4


R.T.: 4.966 min
Delta R.T.: -0.002 min
Response: 11283774
Conc: 534.10 ng/ml



#20 AR-1242-5

R.T.: 6.552 min
Delta R.T.: -0.002 min
Response: 23761522
Conc: 447.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.489 min
Delta R.T.: -0.002 min
Response: 14382223
Conc: 550.20 ng/ml