

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070123\
 Data File : P0096129.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 04:02
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 05:10:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 01 05:09:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.412	3.629	166.6E6	72523371	50.000	50.000
2) SA Decachlor...	10.237	8.662	178.9E6	101.6E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.700	7.553	142.6E6	90373764	500.000	500.000
42) L9 AR-1268-2	8.795	7.618	130.5E6	80601712	500.000	500.000
43) L9 AR-1268-3	9.029	7.825	118.1E6	71771398	500.000	500.000
44) L9 AR-1268-4	9.463	8.114	43654746	29559404	500.000	500.000
45) L9 AR-1268-5	9.889	8.406	374.2E6	203.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070123\
Data File : P0096129.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 04:02
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 05:10:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 01 05:09:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

