

Data Path : Z:\bestpcbsrv\HPCHEM1\ECD 0\Data\P0070218\
 Data File : P0046424.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jul 2018 10:58
 Operator : SM/UA
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 12:03:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 02 11:43:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.385	3.607	14424749	5039134	51.383	48.115
2) SA Decachlor...	10.047	8.566	12414812	4226142	47.200	46.847
Target Compounds						
8) L2 AR-1221-1	4.590	3.819	1931201	793622	500.000	500.000
9) L2 AR-1221-2	4.675	3.903	1425553	572581	500.000	500.000
10) L2 AR-1221-3	4.751	3.979	4575290	1784920	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0070218\
Data File : P0046424.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2018 10:58
Operator : SM/UA
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 12:03:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 02 11:43:49 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.385 min
Delta R.T.: 0.000 min
Response: 14424749
Conc: 51.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.607 min
Delta R.T.: 0.000 min
Response: 5039134
Conc: 48.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.047 min
Delta R.T.: 0.000 min
Response: 12414812
Conc: 47.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.566 min
Delta R.T.: 0.000 min
Response: 4226142
Conc: 46.85 ng/ml



#8 AR-1221-1


R.T.: 4.590 min
Delta R.T.: 0.000 min
Response: 1931201
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.819 min
Delta R.T.: 0.000 min
Response: 793622
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.675 min
Delta R.T.: 0.000 min
Response: 1425553
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.903 min
Delta R.T.: 0.000 min
Response: 572581
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.751 min
Delta R.T.: 0.000 min
Response: 4575290
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.979 min
Delta R.T.: 0.000 min
Response: 1784920
Conc: 500.00 ng/ml