Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070219\
Data File : PO057685.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2019 18:19
Operator : SM/SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED
Ankita
7/3/2019 9:40:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 04:25:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.214	3.547	2101929	1845371	53.509	52.481
2)	SA Decachlor...	9.752	8.462	3031844	2074614	53.916	55.461
	Target Compounds						
3)	L1 AR-1016-1	5.372	4.611	953322	673896	539.374m	504.043m
4)	L1 AR-1016-2	5.395	4.629	1301992	951814	505.239m	488.892m
5)	L1 AR-1016-3	5.455	4.802	804815	524288	516.285	495.201m
6)	L1 AR-1016-4	5.554	4.843	694005	420776	535.172m	472.938m
7)	L1 AR-1016-5	5.843	5.053	697034	550447	509.959m	481.831
31)	L7 AR-1260-1	6.957	6.071	1549993	1127186	559.288	526.650
32)	L7 AR-1260-2	7.213	6.260	2028031	1457351	550.109	531.178
33)	L7 AR-1260-3	7.568	6.410	1722318	1307577	566.899	535.562
34)	L7 AR-1260-4	7.794	6.876	1643693	1148922	541.984	551.825
35)	L7 AR-1260-5	8.100	7.119	3437258	2771531	549.649	546.423

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070219\
Data File : PO057685.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2019 18:19
Operator : SM/SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED
Ankita
7/3/2019 9:40:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 04:25:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062419.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 24 14:02:43 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm