

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
Data File : P0069364.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2020 12:59
Operator : DD\AJ
Sample : L3172-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP2-SS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 13:42:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 29 12:05:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.565	892457	966933	30.168	25.122
2) SA Decachlor...	10.063	8.776	1126631	1148879	23.720	21.105

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
Data File : P0069364.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2020 12:59
Operator : DD\AJ
Sample : L3172-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP2-SS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 13:42:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 29 12:05:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.399 min
Delta R.T.: -0.005 min
Response: 892457
Conc: 30.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP2-SS




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: 0.002 min
Response: 966933
Conc: 25.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.063 min
Delta R.T.: -0.007 min
Response: 1126631
Conc: 23.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.776 min
Delta R.T.: 0.007 min
Response: 1148879
Conc: 21.11 ng/ml