

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
 Data File : P0069376.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jul 2020 16:24
 Operator : DD\AJ
 Sample : L3172-25
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP10-SS

Manual Integrations
 APPROVED

Ankita
 7/6/2020 2:15:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 17:09:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 29 12:05:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.397	3.565	915403	1000847	30.943m	26.003
2) SA Decachlor...	10.060	8.776	1082199	1122715	22.785	20.625

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
Data File : P0069376.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2020 16:24
Operator : DD\AJ
Sample : L3172-25
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
TP10-SS


Manual Integrations
APPROVED

Ankita
7/6/2020 2:15:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 17:09:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 29 12:05:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.397 min
Delta R.T.: -0.007 min
Response: 915403
Conc: 30.94 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TP10-SS


Manual Integrations
APPROVED

Ankita
7/6/2020 2:15:32 PM




#1 Tetrachloro-m-xylene

R.T.: 3.565 min
Delta R.T.: 0.002 min
Response: 1000847
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.060 min
Delta R.T.: -0.009 min
Response: 1082199
Conc: 22.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.776 min
Delta R.T.: 0.006 min
Response: 1122715
Conc: 20.62 ng/ml