

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
 Data File : P0069387.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jul 2020 19:33
 Operator : DD\AJ
 Sample : L3175-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PRESS-SAMPLE

Manual Integrations
 APPROVED

Ankita
 7/6/2020 2:15:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 01:36:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 29 12:05:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.399	3.567	903246	1171194	30.532	30.429
2) SA Decachlor...	10.078	8.784	1471326	4652122	30.978m	85.461 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070220\
Data File : P0069387.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2020 19:33
Operator : DD\AJ
Sample : L3175-01
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PRESS-SAMPLE


Manual Integrations
APPROVED

Ankita
7/6/2020 2:15:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 01:36:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 29 12:05:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.399 min
Delta R.T.: -0.005 min
Response: 903246
Conc: 30.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
PRESS-SAMPLE


Manual Integrations
APPROVED

Ankita
7/6/2020 2:15:38 PM




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: 0.004 min
Response: 1171194
Conc: 30.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.078 min
Delta R.T.: 0.009 min
Response: 1471326
Conc: 30.98 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.784 min
Delta R.T.: 0.014 min
Response: 4652122
Conc: 85.46 ng/ml