

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070224\
 Data File : P0104609.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2024 07:28
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 11:56:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 11:55:46 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.425	3.542	456.0E6	166.7E6	50.000	50.000
2) SA Decachlor...	10.206	8.441	308.3E6	105.5E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.592	4.599	149.1E6	57128916	500.000	500.000
4) L1 AR-1016-2	5.615	4.617	208.9E6	78635455	500.000	500.000
5) L1 AR-1016-3	5.677	4.789	128.4E6	42389185	500.000	500.000
6) L1 AR-1016-4	5.776	4.831	106.3E6	33677003	500.000	500.000
7) L1 AR-1016-5	6.073	5.040	103.8E6	44218824	500.000	500.000
31) L7 AR-1260-1	7.205	6.055	179.4E6	80042104	500.000	500.000
32) L7 AR-1260-2	7.464	6.241	212.0E6	95396419	500.000	500.000
33) L7 AR-1260-3	7.826	6.392	147.9E6	88416251	500.000	500.000
34) L7 AR-1260-4	8.054	6.858	176.0E6	67368037	500.000	500.000
35) L7 AR-1260-5	8.376	7.097	331.4E6	153.9E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070224\
Data File : PO104609.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2024 07:28
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 11:56:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 11:55:46 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.425 min
Delta R.T.: 0.000 min
Response: 455999722
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: 0.000 min
Response: 166675928
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: 0.000 min
Response: 308263994
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.441 min
Delta R.T.: 0.000 min
Response: 105500384
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 149104271
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.599 min
Delta R.T.: 0.000 min
Response: 57128916
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.615 min
Delta R.T.: 0.000 min
Response: 208875436
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.617 min
Delta R.T.: 0.000 min
Response: 78635455
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.677 min
Delta R.T.: 0.000 min
Response: 128429607
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.789 min
Delta R.T.: 0.000 min
Response: 42389185
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.776 min
Delta R.T.: 0.000 min
Response: 106299274
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 33677003
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.073 min
Delta R.T.: 0.000 min
Response: 103832540
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 44218824
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 179359602
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.055 min
Delta R.T.: 0.000 min
Response: 80042104
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.464 min
Delta R.T.: 0.000 min
Response: 211954472
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.241 min
Delta R.T.: 0.000 min
Response: 95396419
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.826 min
Delta R.T.: 0.000 min
Response: 147861144
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 88416251
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.054 min
Delta R.T.: 0.000 min
Response: 175969982
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.858 min
Delta R.T.: 0.000 min
Response: 67368037
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.376 min
Delta R.T.: 0.000 min
Response: 331419581
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.097 min
Delta R.T.: 0.000 min
Response: 153924119
Conc: 500.00 ng/ml