

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070224\
 Data File : P0104624.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2024 11:47
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 12:37:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 12:34:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.542	859.7E6	330.3E6	98.196	98.860
2) SA Decachlor...	10.207	8.440	575.1E6	198.5E6	96.876	96.373
Target Compounds						
26) L6 AR-1254-1	6.453	5.385	280.0E6	157.4E6	958.606	966.802
27) L6 AR-1254-2	6.673	5.530	420.4E6	138.0E6	964.730	962.834
28) L6 AR-1254-3	7.042	5.927	434.5E6	219.7E6	969.572	972.731
29) L6 AR-1254-4	7.328	6.152	322.9E6	136.9E6	968.534	970.436
30) L6 AR-1254-5	7.751	6.565	351.1E6	192.6E6	967.090	969.451



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070224\
Data File : PO104624.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2024 11:47
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 12:37:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 12:34:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: 0.000 min
Response: 859674423
Conc: 98.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.542 min
Delta R.T.: 0.000 min
Response: 330312659
Conc: 98.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: 0.000 min
Response: 575139797
Conc: 96.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.440 min
Delta R.T.: 0.000 min
Response: 198486099
Conc: 96.37 ng/ml



#26 AR-1254-1


R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 280016087
Conc: 958.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.385 min
Delta R.T.: 0.000 min
Response: 157388992
Conc: 966.80 ng/ml




#27 AR-1254-2

R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 420351247
Conc: 964.73 ng/ml



#27 AR-1254-2


R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 137980415
Conc: 962.83 ng/ml



#28 AR-1254-3


R.T.: 7.042 min
Delta R.T.: 0.000 min
Response: 434526017
Conc: 969.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 5.927 min
Delta R.T.: 0.000 min
Response: 219692624
Conc: 972.73 ng/ml




#29 AR-1254-4

R.T.: 7.328 min
Delta R.T.: 0.000 min
Response: 322911658
Conc: 968.53 ng/ml



#29 AR-1254-4


R.T.: 6.152 min
Delta R.T.: 0.000 min
Response: 136928026
Conc: 970.44 ng/ml



#30 AR-1254-5

R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 351068913
Conc: 967.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254\ICC1000



#30 AR-1254-5

R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 192597529
Conc: 969.45 ng/ml