

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070620\
 Data File : P0069394.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Jul 2020 10:50
 Operator : DD\AJ
 Sample : L3173-01DL 2X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 13854DL

Manual Integrations
 APPROVED

Ankita
 7/7/2020 1:12:37 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 07 02:52:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 29 12:05:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.577	293271	297118	9.913	7.719
2) SA Decachlor...	10.065	8.781	426761	487698	8.985	8.959
Target Compounds						
21) L5 AR-1248-1	5.719	4.778	886034	1464008	1136.536	1692.658 #
22) L5 AR-1248-2	6.011	5.021	645226	427078	630.996	354.024m#
23) L5 AR-1248-3	6.222	5.081	613043	555652	495.537	491.098m
24) L5 AR-1248-4	6.651	5.236	762535	181991	507.015	125.856m#
25) L5 AR-1248-5	6.689	5.682	663345	1545221	458.079	1068.095 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070620\
Data File : PO069394.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Jul 2020 10:50
Operator : DD\AJ
Sample : L3173-01DL 2X
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
13854DL


Manual Integrations
APPROVED

Ankita
7/7/2020 1:12:37 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 07 02:52:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061820.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 29 12:05:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.404 min
Delta R.T.: 0.000 min
Response: 293271
Conc: 9.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
13854DL


Manual Integrations
APPROVED

Ankita
7/7/2020 1:12:37 PM




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: 0.014 min
Response: 297118
Conc: 7.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.065 min
Delta R.T.: -0.005 min
Response: 426761
Conc: 8.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.781 min
Delta R.T.: 0.011 min
Response: 487698
Conc: 8.96 ng/ml




#21 AR-1248-1

R.T.: 5.719 min
Delta R.T.: 0.007 min
Response: 886034
Conc: 1136.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
13854DL


Manual Integrations
APPROVED

Ankita
7/7/2020 1:12:37 PM




#21 AR-1248-1

R.T.: 4.778 min
Delta R.T.: -0.016 min
Response: 1464008
Conc: 1692.66 ng/ml




#22 AR-1248-2

R.T.: 6.011 min
Delta R.T.: 0.008 min
Response: 645226
Conc: 631.00 ng/ml



#22 AR-1248-2

R.T.: 5.021 min
Delta R.T.: -0.035 min
Response: 427078
Conc: 354.02 ng/ml m




#23 AR-1248-3

R.T.: 6.222 min
Delta R.T.: 0.008 min
Response: 613043
Conc: 495.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
13854DL


Manual Integrations
APPROVED

Ankita
7/7/2020 1:12:37 PM




#23 AR-1248-3

R.T.: 5.081 min
Delta R.T.: -0.017 min
Response: 555652
Conc: 491.10 ng/ml m




#24 AR-1248-4

R.T.: 6.651 min
Delta R.T.: 0.011 min
Response: 762535
Conc: 507.01 ng/ml



#24 AR-1248-4

R.T.: 5.236 min
Delta R.T.: -0.041 min
Response: 181991
Conc: 125.86 ng/ml m




#25 AR-1248-5

R.T.: 6.689 min
Delta R.T.: 0.012 min
Response: 663345
Conc: 458.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
13854DL

Manual Integrations
APPROVED

Ankita
7/7/2020 1:12:37 PM



#25 AR-1248-5

R.T.: 5.682 min
Delta R.T.: -0.013 min
Response: 1545221
Conc: 1068.09 ng/ml