

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0070718\
 Data File : P0046454.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jul 2018 00:25
 Operator : SM/UA
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 06:22:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 09 06:21:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.212	3.539	11806982	10090432	49.359	48.038
2) SA Decachlor...	9.748	8.472	8987884	9187773	47.570	47.783
Target Compounds						
8) L2 AR-1221-1	4.416	3.751	1591285	1337089	500.000	500.000
9) L2 AR-1221-2	4.500	3.835	1182673	1077988	500.000	500.000
10) L2 AR-1221-3	4.575	3.910	3779366	3425486	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0070718\
Data File : P0046454.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jul 2018 00:25
Operator : SM/UA
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 06:22:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 06:21:24 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.212 min
Delta R.T.: 0.000 min
Response: 11806982
Conc: 49.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 10090432
Conc: 48.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.748 min
Delta R.T.: 0.000 min
Response: 8987884
Conc: 47.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.472 min
Delta R.T.: 0.000 min
Response: 9187773
Conc: 47.78 ng/ml



#8 AR-1221-1


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 1591285
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.751 min
Delta R.T.: 0.000 min
Response: 1337089
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.500 min
Delta R.T.: 0.000 min
Response: 1182673
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.835 min
Delta R.T.: 0.000 min
Response: 1077988
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.575 min
Delta R.T.: 0.000 min
Response: 3779366
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.910 min
Delta R.T.: 0.000 min
Response: 3425486
Conc: 500.00 ng/ml