

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0070822\
 Data File : P0087854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2022 09:29
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 06:18:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.446	3.606	131.0E6	49595830	50.604	50.403
2) SA Decachlor...	10.294	8.609	104.5E6	35217768	51.155	43.443
Target Compounds						
3) L1 AR-1016-1	5.622	4.685	45454651	18910461	506.437	472.873
4) L1 AR-1016-2	5.645	4.704	64129014	26071886	510.260	467.650
5) L1 AR-1016-3	5.708	4.878	40280343	14179699	517.542	481.643
6) L1 AR-1016-4	5.807	4.921	32408973	11558264	530.214	480.973
7) L1 AR-1016-5	6.104	5.134	32516214	14758379	542.560	482.489
31) L7 AR-1260-1	7.237	6.167	56290684	27376391	502.133	484.923
32) L7 AR-1260-2	7.495	6.355	64235142	32117028	471.361	474.710
33) L7 AR-1260-3	7.856	6.509	44607420	29644511	476.229	469.289
34) L7 AR-1260-4	8.083	6.981	52548815	21871483	502.338	470.309
35) L7 AR-1260-5	8.412	7.222	100.4E6	48613650	509.010	450.985



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO070822\
Data File : PO087854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2022 09:29
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 06:18:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.446 min
Delta R.T.: -0.005 min
Response: 130981240
Conc: 50.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: -0.005 min
Response: 49595830
Conc: 50.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.294 min
Delta R.T.: -0.011 min
Response: 104493301
Conc: 51.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.609 min
Delta R.T.: -0.007 min
Response: 35217768
Conc: 43.44 ng/ml



#3 AR-1016-1


R.T.: 5.622 min
Delta R.T.: -0.005 min
Response: 45454651
Conc: 506.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.685 min
Delta R.T.: -0.005 min
Response: 18910461
Conc: 472.87 ng/ml




#4 AR-1016-2

R.T.: 5.645 min
Delta R.T.: -0.005 min
Response: 64129014
Conc: 510.26 ng/ml



#4 AR-1016-2


R.T.: 4.704 min
Delta R.T.: -0.005 min
Response: 26071886
Conc: 467.65 ng/ml



#5 AR-1016-3


R.T.: 5.708 min
Delta R.T.: -0.005 min
Response: 40280343
Conc: 517.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.878 min
Delta R.T.: -0.006 min
Response: 14179699
Conc: 481.64 ng/ml




#6 AR-1016-4

R.T.: 5.807 min
Delta R.T.: -0.006 min
Response: 32408973
Conc: 530.21 ng/ml



#6 AR-1016-4


R.T.: 4.921 min
Delta R.T.: -0.005 min
Response: 11558264
Conc: 480.97 ng/ml



#7 AR-1016-5


R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 32516214
Conc: 542.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.134 min
Delta R.T.: -0.006 min
Response: 14758379
Conc: 482.49 ng/ml




#31 AR-1260-1

R.T.: 7.237 min
Delta R.T.: -0.006 min
Response: 56290684
Conc: 502.13 ng/ml



#31 AR-1260-1


R.T.: 6.167 min
Delta R.T.: -0.006 min
Response: 27376391
Conc: 484.92 ng/ml



#32 AR-1260-2


R.T.: 7.495 min
Delta R.T.: -0.006 min
Response: 64235142
Conc: 471.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.355 min
Delta R.T.: -0.006 min
Response: 32117028
Conc: 474.71 ng/ml




#33 AR-1260-3

R.T.: 7.856 min
Delta R.T.: -0.006 min
Response: 44607420
Conc: 476.23 ng/ml



#33 AR-1260-3


R.T.: 6.509 min
Delta R.T.: -0.006 min
Response: 29644511
Conc: 469.29 ng/ml



#34 AR-1260-4


R.T.: 8.083 min
Delta R.T.: -0.005 min
Response: 52548815
Conc: 502.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.981 min
Delta R.T.: -0.006 min
Response: 21871483
Conc: 470.31 ng/ml



#35 AR-1260-5

R.T.: 8.412 min
Delta R.T.: -0.006 min
Response: 100409713
Conc: 509.01 ng/ml



#35 AR-1260-5

R.T.: 7.222 min
Delta R.T.: -0.006 min
Response: 48613650
Conc: 450.98 ng/ml