

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0071218\
Data File : P0046661.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jul 2018 4:12
Operator : SM/SJ
Sample : J3825-01
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-02-071118-A

Manual Integrations
APPROVED

Sohil
7/13/2018 2:19:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 08:03:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.206	3.531	4292553	5799362	17.962m	27.303 #
2)	SA Decachlor...	9.729	8.458	3145541	2684312	16.503	13.838

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071218\
Data File : P0046661.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Jul 2018 4:12
Operator : SM/SJ
Sample : J3825-01
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CL-02-071118-A


Manual Integrations
APPROVED

Sohil
7/13/2018 2:19:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 08:03:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.206 min
Delta R.T.: -0.006 min
Response: 4292553
Conc: 17.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
CL-02-071118-A


Manual Integrations
APPROVED

Sohil
7/13/2018 2:19:45 PM




#1 Tetrachloro-m-xylene

R.T.: 3.531 min
Delta R.T.: -0.008 min
Response: 5799362
Conc: 27.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.729 min
Delta R.T.: -0.019 min
Response: 3145541
Conc: 16.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.458 min
Delta R.T.: -0.015 min
Response: 2684312
Conc: 13.84 ng/ml