

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071222\
 Data File : P0087952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2022 12:13
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 12:56:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.454	3.609	123.4E6	44579330	47.683	45.305
2)	SA Decachlor...	10.311	8.616	96151206	43343921	47.071	53.467
Target Compounds							
16)	L4 AR-1242-1	5.631	4.689	32331400	13974971	445.870	489.657
17)	L4 AR-1242-2	5.654	4.708	45833533	19496113	448.894	499.778
18)	L4 AR-1242-3	5.717	4.883	28587562	10779307	445.125	517.143
19)	L4 AR-1242-4	5.816	4.967	22978823	10912299	451.376	516.513
20)	L4 AR-1242-5	6.558	5.490	25025510	13159572	471.350	503.428



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071222\
Data File : PO087952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2022 12:13
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 12:56:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.454 min
Delta R.T.: 0.003 min
Response: 123420731
Conc: 47.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.609 min
Delta R.T.: -0.002 min
Response: 44579330
Conc: 45.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.311 min
Delta R.T.: 0.006 min
Response: 96151206
Conc: 47.07 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.616 min
Delta R.T.: 0.000 min
Response: 43343921
Conc: 53.47 ng/ml



#16 AR-1242-1


R.T.: 5.631 min
Delta R.T.: 0.003 min
Response: 32331400
Conc: 445.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.689 min
Delta R.T.: -0.001 min
Response: 13974971
Conc: 489.66 ng/ml




#17 AR-1242-2

R.T.: 5.654 min
Delta R.T.: 0.003 min
Response: 45833533
Conc: 448.89 ng/ml



#17 AR-1242-2


R.T.: 4.708 min
Delta R.T.: -0.001 min
Response: 19496113
Conc: 499.78 ng/ml



#18 AR-1242-3


R.T.: 5.717 min
Delta R.T.: 0.002 min
Response: 28587562
Conc: 445.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.883 min
Delta R.T.: -0.001 min
Response: 10779307
Conc: 517.14 ng/ml




#19 AR-1242-4

R.T.: 5.816 min
Delta R.T.: 0.003 min
Response: 22978823
Conc: 451.38 ng/ml



#19 AR-1242-4


R.T.: 4.967 min
Delta R.T.: -0.001 min
Response: 10912299
Conc: 516.51 ng/ml



#20 AR-1242-5

R.T.: 6.558 min
Delta R.T.: 0.004 min
Response: 25025510
Conc: 471.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.490 min
Delta R.T.: -0.001 min
Response: 13159572
Conc: 503.43 ng/ml