

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071322\
 Data File : P0087981.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 10:50
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 06:35:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.606	126.4E6	45102224	48.844	45.836
2) SA Decachlor...	10.296	8.608	103.3E6	47860049	50.562	59.038
Target Compounds						
16) L4 AR-1242-1	5.622	4.685	33511799	14110783	462.148	494.416
17) L4 AR-1242-2	5.645	4.703	47425558	19834332	464.487	508.448
18) L4 AR-1242-3	5.708	4.879	29854923	10847962	464.858	520.437
19) L4 AR-1242-4	5.807	4.962	23834232	11215422	468.179	530.861
20) L4 AR-1242-5	6.549	5.485	25654637	13419502	483.199	513.372



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071322\
Data File : PO087981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 10:50
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 06:35:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: -0.004 min
Response: 126425762
Conc: 48.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: -0.005 min
Response: 45102224
Conc: 45.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.296 min
Delta R.T.: -0.009 min
Response: 103281121
Conc: 50.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.608 min
Delta R.T.: -0.008 min
Response: 47860049
Conc: 59.04 ng/ml



#16 AR-1242-1


R.T.: 5.622 min
Delta R.T.: -0.005 min
Response: 33511799
Conc: 462.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.685 min
Delta R.T.: -0.006 min
Response: 14110783
Conc: 494.42 ng/ml




#17 AR-1242-2

R.T.: 5.645 min
Delta R.T.: -0.006 min
Response: 47425558
Conc: 464.49 ng/ml



#17 AR-1242-2


R.T.: 4.703 min
Delta R.T.: -0.006 min
Response: 19834332
Conc: 508.45 ng/ml



#18 AR-1242-3


R.T.: 5.708 min
Delta R.T.: -0.006 min
Response: 29854923
Conc: 464.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.879 min
Delta R.T.: -0.005 min
Response: 10847962
Conc: 520.44 ng/ml




#19 AR-1242-4

R.T.: 5.807 min
Delta R.T.: -0.006 min
Response: 23834232
Conc: 468.18 ng/ml



#19 AR-1242-4


R.T.: 4.962 min
Delta R.T.: -0.006 min
Response: 11215422
Conc: 530.86 ng/ml



#20 AR-1242-5

R.T.: 6.549 min
Delta R.T.: -0.006 min
Response: 25654637
Conc: 483.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.485 min
Delta R.T.: -0.006 min
Response: 13419502
Conc: 513.37 ng/ml