

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071322\
 Data File : P0088012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 20:57
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 07:18:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.607	119.9E6	46631691	46.320	47.390
2) SA Decachlor...	10.292	8.607	95539716	47157678	46.772	58.172
Target Compounds						
21) L5 AR-1248-1	5.623	4.686	24548728	11099524	427.620	478.300
22) L5 AR-1248-2	5.899	4.922	35122414	15881967	421.274	505.308
23) L5 AR-1248-3	6.104	4.963	38448291	16353087	429.147	503.821
24) L5 AR-1248-4	6.509	5.134	42088484	19436300	427.882	503.799
25) L5 AR-1248-5	6.549	5.528	40774774	16967643	424.186	465.483



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071322\
Data File : PO088012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 20:57
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 07:18:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: -0.004 min
Response: 119891476
Conc: 46.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.607 min
Delta R.T.: -0.004 min
Response: 46631691
Conc: 47.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.292 min
Delta R.T.: -0.013 min
Response: 95539716
Conc: 46.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.607 min
Delta R.T.: -0.008 min
Response: 47157678
Conc: 58.17 ng/ml



#21 AR-1248-1


R.T.: 5.623 min
Delta R.T.: -0.005 min
Response: 24548728
Conc: 427.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.686 min
Delta R.T.: -0.005 min
Response: 11099524
Conc: 478.30 ng/ml




#22 AR-1248-2

R.T.: 5.899 min
Delta R.T.: -0.005 min
Response: 35122414
Conc: 421.27 ng/ml



#22 AR-1248-2


R.T.: 4.922 min
Delta R.T.: -0.005 min
Response: 15881967
Conc: 505.31 ng/ml



#23 AR-1248-3


R.T.: 6.104 min
Delta R.T.: -0.005 min
Response: 38448291
Conc: 429.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.963 min
Delta R.T.: -0.005 min
Response: 16353087
Conc: 503.82 ng/ml




#24 AR-1248-4

R.T.: 6.509 min
Delta R.T.: -0.006 min
Response: 42088484
Conc: 427.88 ng/ml



#24 AR-1248-4


R.T.: 5.134 min
Delta R.T.: -0.006 min
Response: 19436300
Conc: 503.80 ng/ml



#25 AR-1248-5

R.T.: 6.549 min
Delta R.T.: -0.006 min
Response: 40774774
Conc: 424.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.528 min
Delta R.T.: -0.005 min
Response: 16967643
Conc: 465.48 ng/ml