

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071322\
 Data File : P0088039.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2022 06:15
 Operator : YP/AJ
 Sample : N3684-04 10X
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC247799-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/14/2022
 Supervised By :Ankita Jodhani 07/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 11:08:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 28 02:46:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.446	3.606	4146780	1661472	1.602m	1.689
2) SA Decachlor...	10.291	8.605	4787008	2185580	2.344	2.696

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071322\
Data File : P0088039.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2022 06:15
Operator : YP/AJ
Sample : N3684-04 10X
Misc :
ALS Vial : 46 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC247799-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/14/2022
Supervised By :Ankita Jodhani 07/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 11:08:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 28 02:46:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.446 min
Delta R.T.: -0.005 min
Response: 4146780
Conc: 1.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC247799-1-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/14/2022
Supervised By :Ankita Jodhani 07/14/2022




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: -0.005 min
Response: 1661472
Conc: 1.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.291 min
Delta R.T.: -0.015 min
Response: 4787008
Conc: 2.34 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.605 min
Delta R.T.: -0.010 min
Response: 2185580
Conc: 2.70 ng/ml