

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0071418\
 Data File : P0046730.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Jul 2018 8:54
 Operator : SM/SJ
 Sample : J3958-07
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CS-CONCRETE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 10:03:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 09 15:27:34 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.196	3.522	5333576	5709311	22.318	26.879
2) SA Decachlor...	9.719	8.449	3879415	3858222	20.353	19.890

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0071418\
Data File : P0046730.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Jul 2018 8:54
Operator : SM/SJ
Sample : J3958-07
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CS-CONCRETE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 10:03:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0070718.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 09 15:27:34 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.196 min
Delta R.T.: -0.016 min
Response: 5333576
Conc: 22.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-CONCRETE




#1 Tetrachloro-m-xylene

R.T.: 3.522 min
Delta R.T.: -0.018 min
Response: 5709311
Conc: 26.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.719 min
Delta R.T.: -0.028 min
Response: 3879415
Conc: 20.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.449 min
Delta R.T.: -0.024 min
Response: 3858222
Conc: 19.89 ng/ml