

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071422\
 Data File : P0088066.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2022 20:48
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 03:38:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0071422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 15 03:38:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.605	67495805	25962264	25.846	25.989
2) SA Decachlor...	10.292	8.606	58232409	29729205	26.632	26.587
Target Compounds						
26) L6 AR-1254-1	6.484	5.486	25495625	15373900	270.343	263.728
27) L6 AR-1254-2	6.704	5.634	38517770	13704895	270.053	265.549
28) L6 AR-1254-3	7.072	6.037	38685441	21436585	265.761	260.403
29) L6 AR-1254-4	7.359	6.266	29026190	13311047	267.341	262.120
30) L6 AR-1254-5	7.780	6.684	32117379	19993916	265.600	260.124



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071422\
Data File : PO088066.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2022 20:48
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 03:38:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO071422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 15 03:38:09 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: 0.000 min
Response: 67495805
Conc: 25.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: 0.000 min
Response: 25962264
Conc: 25.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.292 min
Delta R.T.: 0.000 min
Response: 58232409
Conc: 26.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: 0.000 min
Response: 29729205
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 6.484 min
Delta R.T.: 0.000 min
Response: 25495625
Conc: 270.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.486 min
Delta R.T.: 0.000 min
Response: 15373900
Conc: 263.73 ng/ml




#27 AR-1254-2

R.T.: 6.704 min
Delta R.T.: 0.000 min
Response: 38517770
Conc: 270.05 ng/ml



#27 AR-1254-2


R.T.: 5.634 min
Delta R.T.: 0.000 min
Response: 13704895
Conc: 265.55 ng/ml



#28 AR-1254-3


R.T.: 7.072 min
Delta R.T.: 0.000 min
Response: 38685441
Conc: 265.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.037 min
Delta R.T.: 0.000 min
Response: 21436585
Conc: 260.40 ng/ml




#29 AR-1254-4

R.T.: 7.359 min
Delta R.T.: 0.000 min
Response: 29026190
Conc: 267.34 ng/ml



#29 AR-1254-4


R.T.: 6.266 min
Delta R.T.: 0.000 min
Response: 13311047
Conc: 262.12 ng/ml



#30 AR-1254-5

R.T.: 7.780 min
Delta R.T.: 0.000 min
Response: 32117379
Conc: 265.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.684 min
Delta R.T.: 0.000 min
Response: 19993916
Conc: 260.12 ng/ml