

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071423\
 Data File : P0096168.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2023 10:17
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 05:18:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 04 02:14:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.410	0.000	50046	0	0.016	N.D. #
2)	SA Decachlor...	10.246	8.607f	76512	14603	0.038	0.012 #
Target Compounds							
3)	L1 AR-1016-1	5.620	0.000	22010	0	0.251	N.D. #
4)	L1 AR-1016-2	5.620	0.000	22010	0	0.174	N.D. #
5)	L1 AR-1016-3	5.707	0.000	931198	0	11.297	N.D. #
6)	L1 AR-1016-4	5.780	0.000	309670	0	4.741	N.D. #
7)	L1 AR-1016-5	6.084	0.000	63345	0	0.976	N.D. #
8)	L2 AR-1221-1	4.632	3.863	68330	176274	1.885	10.346 #
9)	L2 AR-1221-2	4.727	3.897	300501	239404	11.062	18.756 #
10)	L2 AR-1221-3	4.795	3.978	499969	250284	6.217	6.446
11)	L3 AR-1232-1	4.795	3.978	499969	250284	8.347	8.613
12)	L3 AR-1232-2	5.327	0.000	95350	0	2.670	N.D. #
13)	L3 AR-1232-3	5.620	0.000	22010	0	0.381	N.D. #
14)	L3 AR-1232-4	5.780	0.000	309670	0	10.378	N.D. #
15)	L3 AR-1232-5	5.875	0.000	224840	0	8.649	N.D. #
16)	L4 AR-1242-1	5.620	0.000	22010	0	0.324	N.D. #
17)	L4 AR-1242-2	5.620	0.000	22010	0	0.226	N.D. #
18)	L4 AR-1242-3	5.707	0.000	931198	0	14.225	N.D. #
19)	L4 AR-1242-4	5.780	0.000	309670	0	6.043	N.D. #
20)	L4 AR-1242-5	6.506	5.469f	109746	49559	2.384	1.425 #
21)	L5 AR-1248-1	5.620	0.000	22010	0	0.400	N.D. #
22)	L5 AR-1248-2	5.875	0.000	224840	0	2.531	N.D. #
23)	L5 AR-1248-3	6.084	0.000	63345	0	0.683	N.D. #
24)	L5 AR-1248-4	6.493	0.000	186935	0	2.122	N.D. #
25)	L5 AR-1248-5	6.506	5.588	109746	829159	1.263	16.712 #
26)	L6 AR-1254-1	6.460	5.469f	16321	49559	0.161	0.681 #
27)	L6 AR-1254-2	6.655	5.679	443747	50137	2.993	0.749 #
28)	L6 AR-1254-3	7.021	6.069	5138453	47691	35.973	0.471 #
29)	L6 AR-1254-4	7.368f	0.000	1340898	0	13.933	N.D. #
30)	L6 AR-1254-5	7.754	6.764f	1784071	130440	16.354	1.426 #
31)	L7 AR-1260-1	7.219	6.144f	1274257	136671	10.678	1.916 #
32)	L7 AR-1260-2	7.429f	6.390	232386	206826	1.794	2.457 #
33)	L7 AR-1260-3	7.754	0.000	1784071	0	12.291	N.D. #
34)	L7 AR-1260-4	8.077	0.000	402095	0	4.117	N.D. #
35)	L7 AR-1260-5	8.336f	7.289	1458851	80706	8.472	0.627 #
36)	L8 AR-1262-1	7.852	6.778f	2828788	48072	19.463	1.166 #
37)	L8 AR-1262-2	8.336f	0.000	1458851	0	6.511	N.D. #
38)	L8 AR-1262-3	8.712	7.500f	6233271	124553	38.670	1.949 #
39)	L8 AR-1262-4	8.829f	0.000	309129	0	2.428	N.D. #
40)	L8 AR-1262-5	9.476	0.000	210290	0	2.633	N.D. #
41)	L9 AR-1268-1	8.712	7.500f	6233271	124553	22.413	0.697 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071423\
Data File : P0096168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2023 10:17
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 05:18:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 04 02:14:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.829	0.000	309129	0	1.220	N.D. #
43)	L9 AR-1268-3	9.022	7.793	395107	1677005	1.732	11.825 #
44)	L9 AR-1268-4	9.476	0.000	210290	0	2.505	N.D. #
45)	L9 AR-1268-5	9.845f	0.000	915412	0	1.281	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071423\
Data File : PO096168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2023 10:17
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 05:18:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 04 02:14:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: -0.004 min
Response: 50046
Conc: 0.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.630 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 10.246 min
Delta R.T.: 0.006 min
Response: 76512
Conc: 0.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.607 min
Delta R.T.: -0.057 min
Response: 14603
Conc: 0.01 ng/ml



#3 AR-1016-1


R.T.: 5.620 min
Delta R.T.: 0.026 min
Response: 22010
Conc: 0.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.719 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.620 min
Delta R.T.: 0.003 min
Response: 22010
Conc: 0.17 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.739 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.707 min
Delta R.T.: 0.028 min
Response: 931198
Conc: 11.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.915 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.780 min
Delta R.T.: 0.002 min
Response: 309670
Conc: 4.74 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.957 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.084 min
Delta R.T.: 0.007 min
Response: 63345
Conc: 0.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.171 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.632 min
Delta R.T.: 0.008 min
Response: 68330
Conc: 1.88 ng/ml



#8 AR-1221-1


R.T.: 3.863 min
Delta R.T.: 0.018 min
Response: 176274
Conc: 10.35 ng/ml



#9 AR-1221-2


R.T.: 4.727 min
Delta R.T.: 0.016 min
Response: 300501
Conc: 11.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.897 min
Delta R.T.: -0.033 min
Response: 239404
Conc: 18.76 ng/ml




#10 AR-1221-3

R.T.: 4.795 min
Delta R.T.: 0.007 min
Response: 499969
Conc: 6.22 ng/ml



#10 AR-1221-3


R.T.: 3.978 min
Delta R.T.: -0.029 min
Response: 250284
Conc: 6.45 ng/ml



#11 AR-1232-1


R.T.: 4.795 min
Delta R.T.: 0.008 min
Response: 499969
Conc: 8.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.978 min
Delta R.T.: -0.029 min
Response: 250284
Conc: 8.61 ng/ml




#12 AR-1232-2

R.T.: 5.327 min
Delta R.T.: 0.003 min
Response: 95350
Conc: 2.67 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 4.738 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.620 min
Delta R.T.: 0.003 min
Response: 22010
Conc: 0.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.914 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.780 min
Delta R.T.: 0.002 min
Response: 309670
Conc: 10.38 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.999 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.875 min
Delta R.T.: 0.004 min
Response: 224840
Conc: 8.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.171 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.620 min
Delta R.T.: 0.026 min
Response: 22010
Conc: 0.32 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.719 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.620 min
Delta R.T.: 0.004 min
Response: 22010
Conc: 0.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.737 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.707 min
Delta R.T.: 0.029 min
Response: 931198
Conc: 14.23 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.914 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.780 min
Delta R.T.: 0.002 min
Response: 309670
Conc: 6.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.999 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.506 min
Delta R.T.: -0.016 min
Response: 109746
Conc: 2.38 ng/ml



#20 AR-1242-5


R.T.: 5.469 min
Delta R.T.: -0.055 min
Response: 49559
Conc: 1.42 ng/ml



#21 AR-1248-1


R.T.: 5.620 min
Delta R.T.: 0.026 min
Response: 22010
Conc: 0.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.719 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.875 min
Delta R.T.: 0.004 min
Response: 224840
Conc: 2.53 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.957 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.084 min
Delta R.T.: 0.007 min
Response: 63345
Conc: 0.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.999 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.493 min
Delta R.T.: 0.009 min
Response: 186935
Conc: 2.12 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.171 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.506 min
Delta R.T.: -0.016 min
Response: 109746
Conc: 1.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.588 min
Delta R.T.: 0.022 min
Response: 829159
Conc: 16.71 ng/ml




#26 AR-1254-1

R.T.: 6.460 min
Delta R.T.: 0.005 min
Response: 16321
Conc: 0.16 ng/ml



#26 AR-1254-1


R.T.: 5.469 min
Delta R.T.: -0.055 min
Response: 49559
Conc: 0.68 ng/ml



#27 AR-1254-2


R.T.: 6.655 min
Delta R.T.: -0.021 min
Response: 443747
Conc: 2.99 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.679 min
Delta R.T.: 0.007 min
Response: 50137
Conc: 0.75 ng/ml




#28 AR-1254-3

R.T.: 7.021 min
Delta R.T.: -0.023 min
Response: 5138453
Conc: 35.97 ng/ml



#28 AR-1254-3


R.T.: 6.069 min
Delta R.T.: -0.008 min
Response: 47691
Conc: 0.47 ng/ml



#29 AR-1254-4


R.T.: 7.368 min
Delta R.T.: 0.036 min
Response: 1340898
Conc: 13.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.306 min
Response: 0
Conc: N.D.




#30 AR-1254-5

R.T.: 7.754 min
Delta R.T.: 0.000 min
Response: 1784071
Conc: 16.35 ng/ml



#30 AR-1254-5


R.T.: 6.764 min
Delta R.T.: 0.038 min
Response: 130440
Conc: 1.43 ng/ml



#31 AR-1260-1


R.T.: 7.219 min
Delta R.T.: 0.007 min
Response: 1274257
Conc: 10.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.144 min
Delta R.T.: -0.065 min
Response: 136671
Conc: 1.92 ng/ml




#32 AR-1260-2

R.T.: 7.429 min
Delta R.T.: -0.042 min
Response: 232386
Conc: 1.79 ng/ml



#32 AR-1260-2


R.T.: 6.390 min
Delta R.T.: -0.009 min
Response: 206826
Conc: 2.46 ng/ml



#33 AR-1260-3


R.T.: 7.754 min
Delta R.T.: -0.001 min
Response: 1784071
Conc: 12.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.553 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.077 min
Delta R.T.: 0.018 min
Response: 402095
Conc: 4.12 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 7.027 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.336 min
Delta R.T.: -0.048 min
Response: 1458851
Conc: 8.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.289 min
Delta R.T.: 0.019 min
Response: 80706
Conc: 0.63 ng/ml




#36 AR-1262-1

R.T.: 7.852 min
Delta R.T.: 0.022 min
Response: 2828788
Conc: 19.46 ng/ml



#36 AR-1262-1


R.T.: 6.778 min
Delta R.T.: -0.040 min
Response: 48072
Conc: 1.17 ng/ml



#37 AR-1262-2


R.T.: 8.336 min
Delta R.T.: -0.045 min
Response: 1458851
Conc: 6.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.025 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.712 min
Delta R.T.: 0.010 min
Response: 6233271
Conc: 38.67 ng/ml



#38 AR-1262-3


R.T.: 7.500 min
Delta R.T.: -0.053 min
Response: 124553
Conc: 1.95 ng/ml



#39 AR-1262-4


R.T.: 8.829 min
Delta R.T.: 0.036 min
Response: 309129
Conc: 2.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.617 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.476 min
Delta R.T.: 0.015 min
Response: 210290
Conc: 2.63 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.113 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.712 min
Delta R.T.: 0.013 min
Response: 6233271
Conc: 22.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.500 min
Delta R.T.: -0.054 min
Response: 124553
Conc: 0.70 ng/ml




#42 AR-1268-2

R.T.: 8.829 min
Delta R.T.: 0.034 min
Response: 309129
Conc: 1.22 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 7.618 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.022 min
Delta R.T.: -0.006 min
Response: 395107
Conc: 1.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.793 min
Delta R.T.: -0.032 min
Response: 1677005
Conc: 11.83 ng/ml




#44 AR-1268-4

R.T.: 9.476 min
Delta R.T.: 0.014 min
Response: 210290
Conc: 2.51 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.114 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.845 min
Delta R.T.: -0.043 min
Response: 915412
Conc: 1.28 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.407 min
Response: 0
Conc: N.D.