

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0071521\
 Data File : P0079586.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Jul 2021 17:56
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:37:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0070821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 10 00:52:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.018	4.227	1448345	566971	21.663	19.771
2)	SA Decachlor...	11.145	9.695f	1151649	506666	22.643	19.145
Target Compounds							
8)	L2 AR-1221-1	5.271	0.000	53973	0	76.170	N.D. #
9)	L2 AR-1221-2	5.373	4.566f	15791	3856	30.459	15.156 #
28)	L6 AR-1254-3	7.894	0.000	22262	0	5.919	N.D. #
32)	L7 AR-1260-2	8.315	0.000	5318	0	1.688	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO071521\
Data File : PO079586.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Jul 2021 17:56
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:37:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO070821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 10 00:52:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.018 min
Delta R.T.: -0.002 min
Response: 1448345
Conc: 21.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.227 min
Delta R.T.: -0.001 min
Response: 566971
Conc: 19.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.145 min
Delta R.T.: -0.005 min
Response: 1151649
Conc: 22.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.695 min
Delta R.T.: 0.018 min
Response: 506666
Conc: 19.14 ng/ml



#8 AR-1221-1


R.T.: 5.271 min
Delta R.T.: -0.002 min
Response: 53973
Conc: 76.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.492 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.373 min
Delta R.T.: 0.000 min
Response: 15791
Conc: 30.46 ng/ml



#9 AR-1221-2


R.T.: 4.566 min
Delta R.T.: -0.027 min
Response: 3856
Conc: 15.16 ng/ml



#28 AR-1254-3


R.T.: 7.894 min
Delta R.T.: 0.014 min
Response: 22262
Conc: 5.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.990 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 8.315 min
Delta R.T.: 0.000 min
Response: 5318
Conc: 1.69 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.327 min
Response: 0
Conc: N.D.